



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 20/2/12

Product Name	CLEARKLENS DE VH29
Product Code	7515783
Batch Number	4031582
Production Date	02/2012
Expiration Date	02/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear colourless liquid
Specific Gravity (20°C)	JDM004	0.880	0.888	0.885

On behalf of Diversey site Quality Manager	Name :	
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7515783

Version : 03

Date of issuing : November 28th 2011

0520278 00

Page 1

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@synergyhealthplc.com
Website: www.synergyhealthplc.com

Cust. Ref: F018143
Date Rec'd 9/02/12
Date: 13/02/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B212020391	JD DE 900ML BOTTLES 25-45kGy	98	B/No.W031582 - P/REF.7515783 ROUTINE DOSE 37.2, 37.1 kGy
B2FLE0010999 B212020391	SAMPLES	5	CUSTOMER BATCH/CODE No.s AS ORDER No.P018143 DATED 8/2/12
Despatch Text: 2PLTS			
Total		103	Last Page 1 of 1

All in accordance with current
Technical Agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of SYNERGY HEALTH PLC



Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 20 Feb 2012
Date Tested: 24 Feb 2012
Date Test Completed: 09 Mar 2012
Purchase Order: 18195

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0003645/4
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 20/02/2012
Test Article: ClearKlens DE Block 2
Sample Code: 3258
Batch Ref: Works Order No. 31582
Qty Received: 20 x 900mL

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 13 Mar 2012 10:37:50

Mrs P. Pham-Lengoc, BSc, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0003645

Certificate of Analysis - OSMM

Print Number: P0001960

Page 1 of 1



Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 20 Feb 2012
Date Tested: 23 Feb 2012
Date Test Completed: 23 Feb 2012
Purchase Order: 18195

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0003645/5
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 20/02/2012
Test Article: ClearKlens DE Block 2
Sample Code: 3258
Batch Ref: Works Order No. 31582
Qty Received: 1 x 900mL

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2500 EU/mL
Spiked Recovery	MM110/01	90 %
Dilution Tested	MM110/04	1/50
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM01

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.



Date: 24 Feb 2012 13:34:32

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Senior Laboratory Manager - Pharmaceutical Microbiology