



Diversey Europe BV

Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands

Tel. +31 (0)30 247 6427

## CERTIFICATE OF ANALYSIS

Date: 20/4/12

Product Name	CLEARKLENS IPA VH1
Product Code	7513400
Batch Number	W031747
Production Date	03/2012
Expiration Date	03/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.877

On behalf of Diversey site Quality Manager	Name :	C. Piscoe
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513400

Version : 03

Date of issuing : November 28<sup>th</sup> 2011

## 0527217 00

Page 1



trial Estate  
synergyhealth  
our work protects your world

Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom

Telephone: +44 (0)1274 686011

Facsimile: +44 (0)1274 686061

Email: [assistance@synergyhealthplc.com](mailto:assistance@synergyhealthplc.com)

Website: [www.synergyhealthplc.com](http://www.synergyhealthplc.com)

LA1 4XS

P018403

Cust. Ref: PO18403  
Date Rec'd 5/04/12  
Date: 13/04/12

All in accordance with current  
Technical Agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: 

For and on behalf of SYNERGY HEALTH PLC



Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

**Date Received:** 24 Apr 2012  
**Date Tested:** 26 Apr 2012  
**Date Test Completed:** 26 Apr 2012  
**Purchase Order:** 18470

### **CERTIFICATE OF ANALYSIS**

**Laboratory Reference Number:** 0004406/6  
**Test Required:** Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP  
**Date Received:** 24/04/2012  
**Test Article:** Clearklens IPA Block1  
**Sample Code:** -  
**Batch Ref:** WO31747  
**Qty Received:** 1 x 900mL

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	87 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

### **Comments**

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 27 Apr 2012 11:45:42

Mrs P. Pham-Lengoc, BSc, MSB  
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0004406

Certificate of Analysis - OSMM

Print Number: P0002709

Page 1 of 1



Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

**Date Received:** 24 Apr 2012  
**Date Tested:** 03 May 2012  
**Date Test Completed:** 17 May 2012  
**Purchase Order:** 18470

### CERTIFICATE OF ANALYSIS

**Laboratory Reference Number:** 0004863/3  
**Test Required:** Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP  
**Date Received:** 24/04/2012  
**Test Article:** Clearklens IPA Block 1  
**Sample Code:** -  
**Batch Ref:** WO31747  
**Qty Received:** 20 x 900mL + 3 Spares

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

### Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 24 May 2012 14:38:07

Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0004863

Certificate of Analysis - OSMM

Print Number: P0003131

Page 1 of 1