

**Diversey Europe BV** 

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

### **CERTIFICATE OF ANALYSIS**

Date: 20/4/12

Product Name	CLEARKLENS IPA VH1	
Product Code	7513400	
Batch Number	WO 31960	
Production Date	04/2012	
Expiration Date	04/2014	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
Appearance	Visual	Clear Co	Upper	clear colonda
	v Iodai	Liquid		clear colourte
Specific Gravity (20°C)	JDM004	0.872	0.883	0.877

On behalf of Diversey site Quality Manager	Name:	Chascoe
	Position	Quality Control Inspector

This document being issued elect	ronically does not bear a signature	
COA 7513400	Version: 03	Date of issuing: November 28th 2011

# CERTIFICATE OF IRRADIATION Number: B

0527716 00

FLE001

Flexible Medical Packaging Ltd

Unit 8

White Cross Industrial Estate

Hightown Lancaster synergy health
our work protects your world

LA1 4XS

Cust. Ref: Date Rec'd P018430 12/04/12

Date: 19/04/12

Page

1

Synergy Health plc

Roydsdale Way · Euroway Trading Estate

Bradford · West Yorkshire BD4 6SE · United Kingdom

This is to certify that the above items have been irradiated as

For and on behalf of SYNERGY HEALTH PLC

specified above

Authorising Signature:

Telephone: +44 (0)1274 686011 Facsimile: +44 (0)1274 686061

Email: assistance@synergyhealthplc.com

Website: www.synergyhealthplc.com

Date: 19/04/	12		
ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B212040384	JD IPA BOTTLES 25-45kGy	530	B/No.W031960
8212040364			ROUTINE DOSE 39.4, 38.4, 41.9 kGy
B2FLE0010999 B212040384	SAMPLES	5	
			CUSTOMER BATCH/CODE No.s AS ORDER No.P01843D DATED 11/4/12
	Despatch Text:		
	10PLTS		
	Total	535	Last Page 1 of 1
		All in acco Technical A	ordance with current agreement



Wickham Laboratories Limited Winchester Road, Wickham, Fareham Hampshire PO17 5EU England Telephone: +44 (0)1329 832511

Fax: +44 (0)1329 834262 mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Mr J Jackson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster

LA14XS

**Date Received:** 

24 Apr 2012

Date Tested:

26 Apr 2012

**Date Test Completed:** 

26 Apr 2012

**Purchase Order:** 

18470

### **CERTIFICATE OF ANALYSIS**

**Laboratory Reference Number:** 

0004406/5

**Test Required: Date Received:**  Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

24/04/2012

Test Article:

Clearklens IPA Block1

Sample Code:

Batch Ref: **Qty Received:** 

WO31960

1 x 900mL

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	119 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

#### Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 27 Apr 2012 11:45:42

Mrs P. Pham-Lengoc, BSc, MSB

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0004406

Certificate of Analysis - OSMM

Print Number: P0002709

Page 1 of 1



Wickham Laboratories Limited Winchester Road, Wickham, Fareham Hampshire PO17 5EU England Telephone: +44 (0)1329 832511

Fax: +44 (0)1329 834262 mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Mr J Jackson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

24 Apr 2012

Date Tested:

03 May 2012

Date Test Completed:

17 May 2012

**Purchase Order:** 

18470

## **CERTIFICATE OF ANALYSIS**

**Laboratory Reference Number:** 

0004863/2

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Date Received:

24/04/2012

Test Article:

Clearklens IPA Block 1

Sample Code:

-

Batch Ref:

WO31960

**Qty Received:** 

20 x 900mL + 3 Spares

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 5 mL
Product Standard Data Sheet	FG047/psd	FM03
Campuanta		

#### **Comments**

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Eur Uses.

Date: 24 May 2012 14:38:07

Mrs S.L. Wood, BSc (Hons), CBiol, MSB

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0004863

Certificate of Analysis - OSMM

Print Number: P0003131

Page 1 of 1