



Diversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 21.05.12

Product Name	CLEARKLENS IPA VH1
Product Code	7514372
Batch Number	10032040
Production Date	05/2012
Expiration Date	05/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		clear colourless Liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.877

On behalf of Diversey site Quality Manager	Name :	Molero
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7514372	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Synergy Health plc

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom

Telephone: +44 (0)1274 686011

Facsimile: +44 (0)1274 686061

Email: assistance@synergyhealthplc.com

Website: www.synergyhealthplc.com

LA1 4XS

P018559

Cust. Ref: 10/05/12

Date Rec'd 17/05/12

Date:

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B212050414	JD IPA BOTTLES 25-45kGy	697	B/No.W032040 - P/REF.7514372 ROUTINE DOSE 38.0, 36.0, 35.6, 35.7 kGy
B2FLE0010999 B212050414	SAMPLES	8	 CUSTOMER BATCH/CODE No.s AS ORDER No.P018559 DATED 9/5/12
Despatch Text: 13PLTS			
Total		705	Last Page 1 of 1

All in accordance with current
Technical Agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature: *J. Coffey*

For and on behalf of SYNERGY HEALTH PLC



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 21 May 2012
Date Tested: 24 May 2012
Date Test Completed: 24 May 2012
Purchase Order: 18603

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0004808/2
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 21/05/2012
Test Article: ClearKlens IPA Block 2
Sample Code: 3460
Batch Ref: WO32040
Qty Received: 1 x 900mL Bottles

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	87 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 30 May 2012 11:51:23

Mrs P. Pham-Lengoc, BSc, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0004808

Certificate of Analysis - OSMM

Print Number: P0003201

Page 1 of 1



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 19 May 2012
Date Tested: 23 May 2012
Date Test Completed: 06 Jun 2012
Purchase Order: 18603

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0004808/1
Test Required: Sterility Test by Membrane Filtration (Steritest)
Date Received: 19/05/2012
Test Article: ClearKlens IPA Block 2
Sample Code: 3460
Batch Ref: WO32040
Qty Received: 20 x 900mL Bottles + Spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Approval is provided by Electronic Signature. Their name is shown below.

Date: 08 Jun 2012 12:10:15

Mrs P. Pham-Lengoc, BSc, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0004808

Certificate of Analysis - OSMM

Print Number: P0003383

Page 1 of 1