



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS IPA VH01
PRODUCT CODE	7513400
BATCH NUMBER	W027859
PRODUCTION DATE	05-2010
EXPIRY DATE	05-2012

TEST	METHOD	LIMITS	RESULTS
Appearance	VISUAL	Clear Colourless Liquid	CLEAR LIQUID
Density (20°C)	GM 0110 01	0.865 – 0.885	0.870

Date: 24/05/2010

QA Manager: [Signature]

QA: E. Gargan

JD Approval: [Signature]

This document has been produced on behalf of JohnsonDiversey by authorised personnel and is valid unsigned.

CERTIFICATE OF IRRADIATIONNumber: **B**

0447406 00

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster

**Isotron Limited**

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P015419
Date Rec'd 13/05/10
Date: 18/05/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B210050328	JD IPA BOTTLES 25-45kGy	353	BATCH NO: W027859-CUSTB/NO: 1699 ACTUAL DOSE....37.4 - 38.2....kGy
B2FLE0010999 B210050328	SAMPLES	5	CUSTOMER BATCH/CODE NO'S ON ORDER NO P015419 DATED 12.05.10
Despatch Text: 7PLTS			
Total		358	Last Page 1 of 1

All in accordance with current
technical agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature:

A handwritten signature in black ink, consisting of a large, stylized 'Q' or 'G' shape followed by a horizontal line.

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 24-May-10
Date Tested : 02-Jun-10
Date Completed : 02-Jun-10
Purchase Order No. : 15461

Test Article : ClearKlens IPA Block 1

Sample Code :

Batch Reference : 1699 W027859

Laboratory Reference : LR-00964487

Article Quantity : 1 x 900mLs

Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2000 EU/mL
Dilution	*	MM110/01	1/40
Spiked Recoveries	*	MM110/02	Valid 71%
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM02

Test carried out according to current Pharmacopoeias.

GLP

Compliant
Laboratory

Suz Wood

Date *10 June 2010*

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN\00150854

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 24-May-10
Date Tested : 02-Jun-10
Date Completed : 16-Jun-10
Purchase Order No. : 15461

Test Article : ClearKlens IPA Block 1

Sample Code :

Batch Reference : 1699 WO27859

Laboratory Reference : LR-00964486

Article Quantity : 20 x 900mLs

Test Required : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 900002

Volume Tested: 20 x 50mL

Test carried out according to current Pharmacopoeias.

GLP

Compliant
Laboratory

Suz Wood

Date *18 June 2010*

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN\00150854

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited