

Goods in



Diversey Europe Operations BV
 Maarssebroeksedijk 2
 3542 DN Utrecht
 The Netherlands
 Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 8.05.2017

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26
Product Code	7515828
Batch Number	FHP 17128 42460
Production Date	08/05/2017
Expiration Date	Exp 08/05/2019

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clean Colourless Liquid
pH (100% 20°C)	JDM001	1.5	2.5	1.7
Specific Gravity (20°C)	JDM004	1.010	1.030	1.016

On behalf of Diversey site Quality Manager	Name :	<i>Alpa Krishna Parthasarathy</i>
	Position	Quality Control Inspector

Goods in.



Diversey Europe Operations BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 8.05.2017

Product Name	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26
Product Code	7515828
Batch Number	MP17128 42460
Production Date	08/05/2017
Expiration Date	Exp 08/05/2019

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
pH (100% 20°C)	JDM001	9.0	12.5	11.7
Specific Gravity (20°C)	JDM004	1.004	1.020	1.008
Potential Chlorine Dioxide (ppm)		100	180	141.75

On behalf of Diversey site Quality Manager	Name :	Olpa Linenna, Poutyisker
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Bi-Spore Activator	Version : 01	Date of issuing : October 1 st 2012
------------------------	--------------	--

Certificate of Irradiation

Date Issued: 18-Apr-2017

UK33S11853720-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
Unit 8, White Cross Ind Estate
Hightown
Lancaster
Lancashire LA1 4XS
UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1106825
Customer Reference Number:	PO28742
Product Description:	BI-SP B3 COMPONENTS DV4725 25-40kGy
Validation Reference:	4725
Quantity Received:	8
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	40.0
Customer Unit Lot/Batch Number:	WO42465, 1 plt

Irradiation Data

Date and Time of Irradiation:	18-Apr-2017 19:11
Reference Dose Range kGy:	33.0 - 33.3
Calculated Minimum Dose kGy:	27.3
Calculated Maximum Dose kGy:	34.0

Product irradiated under WO42465 will be used in finished product batch FMP17128 42460

Irradiation Release Authorised By Synergy Health plc

Processing Site: Brunel Close, Drayton Fields Industrial Estate, Daventry, NN11 8RB Phone No: +44 (0) 1327 706111

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069



Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 15 May 2017
Date Tested: 24 May 2017
Date Test Completed: 07 Jun 2017
Purchase Order: 28905

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0035577/1
Test Required: Sterility by Membrane Filtration Steritest
Date Received: 15/05/2017
Test Article: Bi-Spore Activator
Sample Code: FMP17128 42460
Batch Ref: 563611
Qty Received: 20 Bottles (100mL)

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	EP 9.0 2.6.1, USP 40 <71> & JP XVII 4.06
Product Standard Data Sheet	FG047/psd	FM04

Approval is provided by Electronic Signature. Their name and position is shown below.

 CB: MKSB

Date: 08 Jun 2017 13:16:42

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology



Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 15 May 2017
Date Tested: 24 May 2017
Date Test Completed: 07 Jun 2017
Purchase Order: 28905

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0035577/2
Test Required: Sterility by Membrane Filtration Steritest
Date Received: 15/05/2017
Test Article: Bi-Spore Base
Sample Code: FMP17128 42460
Batch Ref: 552908
Qty Received: 20 Bottles (100mL)

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	EP 9.0 2.6.1, USP 40 <71> & JP XVII 4.06
Product Standard Data Sheet	FG047/psd	FM05

Approval is provided by Electronic Signature. Their name and position is shown below.

CBio: MRSB

Date: 08 Jun 2017 13:16:42

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology