



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 12/11/12

Product Name	CLEARKLENS TEGO 2000SC VH25
Product Code	7513387
Batch Number	FMP12299 32940
Production Date	10/2012
Expiration Date	10/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	Clear slightly yellow liquid
pH 100% (20°C)	JDM001	7.5	8.5	7.9
Specific Gravity (20°C)	JDM004	0.993	1.003	1.001

On behalf of Diversey site Quality Manager	Name :	
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513387	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--

CERTIFICATE OF IRRADIATION

Number: **B**

0551637 00

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Synergy Health plc

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom

Telephone: +44 (0)1274 686011

Facsimile: +44 (0)1274 686061

Email: assistance@synergyhealthplc.com

Website: www.synergyhealthplc.com

LA1 4XS

Cust. Ref: P019384

Date Rec'd 1/11/12

Date: 5/11/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B212110232	DIVERSEY TEGO 2000SC 50ml 25-42kGy	104	B/No.W032940 CUSTOMER B/No.FMP12299 32940 ROUTINE DOSE 33.3, 32.5, 31.6kGy
B2FLE0010999 B212110232	SAMPLES	1	CUSTOMER BATCH/CODE no.5 AS ORDER No.P019384 DATED 31/10/12
Despatch Text: 2PLTS			
Total		105	Last Page 1 of 1

All in accordance with current
Technical Agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature:

For and on behalf of SYNERGY HEALTH PLC



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 07 Nov 2012
Date Tested: 20 Nov 2012
Date Test Completed: 12 Dec 2012
Purchase Order: PO19417

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0007553/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I with Subculture: PH Eur and USP
Date Received: 07/11/2012
Test Article: Tego 2000sc Block 1
Sample Code: 410594
Batch Ref: FMP12299 32940
Qty Received: 20 x 100mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Growth of Sub-culture in Tryptone Soya Broth at 30-35°C after 4 days	MM107/07	No growth in two broths
Growth of Sub-culture in Fluid Thioglycollate at 20-25°C after 4 days	MM107/08	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM08

Comments

As Steritest Sterility Test

Approval is provided by Electronic Signature. Their name is shown below.

Suz Wood

Date: 12 Dec 2012 16:13:41

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0007553

Certificate of Analysis - OSMM

Print Number: P0007469

Page 1 of 1