



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 28/11/12

Product Name	CLEARKLENS IPA VH1
Product Code	7513400
Batch Number	FMP12304 32859
Production Date	10/2012
Expiration Date	10/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		clear colourless liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.877

On behalf of Diversey site Quality Manager	Name :	
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513400	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--

CERTIFICATE OF IRRADIATION

Number: **B**

0552378 00

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Synergy Health plc

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom

Telephone: +44 (0)1274 686011

Facsimile: +44 (0)1274 686061

Email: assistance@synergyhealthplc.com

Website: www.synergyhealthplc.com

LA1 4XS

P019440

Cust. Ref: 8/11/12

Date Rec'd 23/11/12

Date:

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B212110453	JD IPA BOTTLES 25-45kGy	709	B/No.W032859 FMP 12304 - 32859 (ca) 26.11.12 ROUTINE DOSE 37.9, 38.8, 38.2, 38.1 kGy's
B2FLE0010999 B212110453	SAMPLES	5	CUSTOMER BATCH/CODE No.s AS ORDER No.P019440 DATED 7/11/12
Despatch Text: 14PLTS			
Total		714	Last Page 1 of 1

All in accordance with current
Technical Agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature:

For and on behalf of SYNERGY HEALTH PLC



Wickham Laboratories
contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 04 Dec 2012
Date Tested: 10 Dec 2012
Date Test Completed: 10 Dec 2012
Purchase Order: PO19549

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0007931/2
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 04/12/2012
Test Article: Clearklens IPA Block 1
Sample Code:
Batch Ref: FMP12304 32859 (3906)
Qty Received: 1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	86 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 14 Dec 2012 10:03:14

Mrs P. Pham-Lengoc, BSc, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0007931

Certificate of Analysis - OSMM

Print Number: P0007538

Page 1 of 1



Wickham Laboratories
contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 28 Nov 2012
Date Tested: 06 Dec 2012
Date Test Completed: 20 Dec 2012
Purchase Order: PO19549

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0007931/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 28/11/2012
Test Article: Clearklens IPA Block 1
Sample Code: 3906
Batch Ref: FMP1230432859
Qty Received: 20 x 900mL Bottles (+3 spare)

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

3 boxes received 28 Nov 2012; 1 box received 04 Dec 2012.
Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Susan Wood

Date: 21 Dec 2012 12:00:13

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0007931

Certificate of Analysis - OSMM

Print Number: P0007685

Page 1 of 1