



CERTIFICATE OF ANALYSIS

Mr J Jackson

Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 16-Sep-08
Date Tested : 24-Sep-08
Date Completed : 08-Oct-08
Purchase Order No. : PO13149

Test Article : Clearklens DE

Sample Code :

Batch Reference : WO24955

Laboratory Reference : LR-00789659

Article Quantity : 20 x 900mL

Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 226049

Volume Tested: 20 x 50mL

Test carried out according to current Pharmacopoeias.

GLP

Compliant
Laboratory

WAB

Date 14 Oct 2008

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00126309

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 16-Sep-08
Date Tested : 01-Oct-08
Date Completed : 01-Oct-08
Purchase Order No. : PO13149

Test Article : Clearklens DE

Sample Code :

Batch Reference : WO24955

Laboratory Reference : LR-00789660

Article Quantity : 1 x 900mL

Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.


Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.25 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	82% Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL

This sample was tested at a 1/50 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.
Product Standard Data Sheet: FM01

GLP

Compliant
Laboratory

Date 7 Oct 2008


K.A. Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00126309

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited

Number: **B** 0382203 00

Page 1



Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

Cust. Ref: P013111
Date Rec'd 5/09/08
Date: 11/09/08

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: C. Drice

For and on behalf of ISOTRON LIMITED

JohnsonDiversey

Clean is just the beginning



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEAR KLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	W024955
PRODUCTION DATE	08/2008
EXPIRY DATE	09/2010

TEST	METHOD	LIMITS	RESULTS
Appearance		Clear Colourless Liquid	clear colourless liquid
Density (20°C)	GM 0110 01	0.876 – 0.896	0.882

Date:.....12/09/08.....

QA Manager:..........

QA:..........

JD Approval:..........

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

Issue no. 4
Issued by W. Ashton
Issue date 07.04.2006