



### CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARLENs IPA VH01
PRODUCT CODE	7513400
BATCH NUMBER	W024957
PRODUCTION DATE	09/2008
EXPIRY DATE	09/2010

TEST	METHOD	LIMITS	RESULTS
Appearance		Clear Colourless Liquid	Clear colourless liquid
Density (20°C)	GM 0110 01	0.865 – 0.885	0.877

Date:.....10/10/08.....

QA Manager:.....*[Signature]*.....

QA:.....*E. Gough*.....

JD Approval:.....*[Signature]*.....

*This document has been produced on behalf of JohnsonDiversey by authorised personnel and is valid unsigned.*

# CERTIFICATE OF IRRADIATION

Number: **B** 0382851 00

FLE001

Page 1

Flexible Medical Packaging Ltd  
Unit 8  
White Cross Industrial Estate  
Hightown  
Lancaster



## Isotron Limited

Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom  
Telephone: +44 (0)1274 686011  
Facsimile: +44 (0)1274 686061  
Email: assistance@isotron.com  
Website: www.isotron.com

LA1 4XS

Cust. Ref: P013143  
Date Rec'd 11/09/08  
Date: 15/09/08

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B208090319	JD IPA BOTTLES 25-45kGy	171	B/No.W024957  ACTUAL....36.0 - 38.4....kGy DOSE
B2FLE0010999 B208090319	SAMPLES	4	  CUSTOMERS BATCH/CODE No.s AS ORDER No.P013143 DATED 10/9/08
Total		175	Last Page 1 of 1

All in accordance with current  
technical agreement

This is to certify that the above items have been irradiated as  
specified above

Authorising Signature: *[Signature]*

For and on behalf of ISOTRON LIMITED



## CERTIFICATE OF ANALYSIS

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

**Date Received** : 23-Sep-08  
**Date Tested** : 25-Sep-08  
**Date Completed** : 09-Oct-08  
**Purchase Order No.** : PO13175


**Test Article** : ClearKlens IPA  
**Sample Code** :  
**Batch Reference** : WO24957  
**Laboratory Reference** : LR-00791345 **Article Quantity** : 20 x 900mL  
**Test Required** : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 226052  
Volume Tested: 1 Litre  
Test carried out according to current Pharmacopoeias.

**GLP**  
Compliant  
Laboratory

Date 14 Oct 2008

  
K.A. Barker, C.Biol, M.I. Biol.,  
Business Manager - Microbiology

**Sample Booking Number:** CN\00126546  
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*



## CERTIFICATE OF ANALYSIS

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

**Date Received** : 23-Sep-08  
**Date Tested** : 01-Oct-08  
**Date Completed** : 01-Oct-08  
**Purchase Order No.** : PO13175

**Test Article** : ClearKlens IPA

**Sample Code** :

**Batch Reference** : WO24957

**Laboratory Reference** : LR-00791346

**Article Quantity** : 20 x 900mL

**Test Required** : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2 EU/mL
Dilution	*	MM110/01	1/40
Spiked Recoveries	*	MM110/02	94% Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL

This sample was tested at a 1/40 dilution and spiked with endotoxin to check for interference.  
Test carried out according to current Pharmacopoeias.  
Product Standard Data Sheet: FM02

**GLP**

Compliant  
Laboratory

Date 7 Oct 2008

W.A.B.  
K.A.Barker, C.Biol, M.I. Biol.,  
Business Manager - Microbiology

**Sample Booking Number:** CN\00126546  
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*