



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	20024964
PRODUCTION DATE	10/2008
EXPIRY DATE	10/2010

TEST	METHOD	LIMITS	RESULTS
Appearance		Clear Colourless Liquid	Clear colourless liquid
Density (20°C)	GM 0110 01	0.876 - 0.896	0.885

Date:..... 03/11/08

QA.Manager:..... 

QA:..... E. Gauran

JD Approval:..... 

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013279
 Date Rec'd 16/10/08
 Date: 24/10/08

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
BZFLE0010121 B208100476	JD DE 900ML BOTTLES 25-45kGy	686	515 X B/No.W024963 - CB/No.F/C14 171 X B/No.W024964 - CB/No.F/C15 ACTUAL...39.4 - 41.6...kGy DOSE
BZFLE0010999 B208100476	SAMPLES	9	CUSTOMERS BATCH/CODE No.s AS ORDER No.P013279 DATED 15/10/08
Total		695	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 31-Oct-08
Date Tested : 11-Nov-08
Date Completed : 11-Nov-08
Purchase Order No. : 13339


Test Article : Clearlens DE Block 1
Sample Code : Works Order No. 24964
Batch Reference : 15
Laboratory Reference : LR-00803117 **Article Quantity** : 2 Bottles
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	1.5887 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	119% Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL

This sample was tested at a 1/50 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.
Product Standard Data Sheet: FM01

W.A.B.

Date 3 Dec 2008



K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN00128282
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 31-Oct-08
Date Tested : 11-Nov-08
Date Completed : 25-Nov-08
Purchase Order No. : 13339

Test Article : Clearklens DE Block 1
Sample Code : Works Order No. 24964
Batch Reference : 15
Laboratory Reference : LR-00803116 **Article Quantity** : 20 Bottles
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 306021
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.



W.A.B.

Date 2 Dec 2008

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00128282
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited