

JohnsonDiversey

Clean is just the beginning



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	W025089
PRODUCTION DATE	11/2008
EXPIRY DATE	11/2010

TEST	METHOD	LIMITS	RESULTS
Appearance		Clear Colourless Liquid	clear colourless liquid
Density (20°C)	GM 0110 01	0.876 – 0.896	0.885

Date:.....19/11/08.....
QA Manager:.....*[Signature]*.....
QA:.....*[Signature]*.....
JD Approval:.....*[Signature]*.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

Issue no. 4
Issued by W. Ashton
Issue date 07.04.2006

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013404
 Date Rec'd 13/11/08
 Date: 18/11/08

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
BZFLE0010121 B208110429	JD DE 900ML BOTTLES 25-45kGy	706	B/No.W025089 CUSTOMER B/No.F/C 146 ACTUAL....36.0 - 38.3....kGy DOSE
BZFLE0010999 B208110429	SAMPLES	10	CUSTOMERS BATCH/CODE No.s AS ORDER No.P013404 DATED 13/11/08
Total		716	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *[Signature]*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 21-Nov-08
Date Tested : 08-Dec-08
Date Completed : 08-Dec-08
Purchase Order No. : PO13435

Test Article : Clearklens DE Block 1
Sample Code : Works Order No. 25089
Batch Reference : 146
Laboratory Reference : LR-00809062 **Article Quantity** : 4 Bottles
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2500 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	66% Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL

This sample was tested at a 1/50 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.
Product Standard Data Sheet: FM01



WKB

Date 16 Dec 2008

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN00129067
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited

CERTIFICATE OF ANALYSIS


Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 21-Nov-08
Date Tested : 26-Nov-08
Date Completed : 10-Dec-08
Purchase Order No. : PO13435

Test Article : Clearlens DE Block 1
Sample Code : Works Order No. 25089
Batch Reference : 146
Laboratory Reference : LR-00809061 **Article Quantity** : 20 Bottles
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM03

Day Book Ref: 306032
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.



WAB

Date 16 Dec 2008

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00129067
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited