



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS IPA VH01
PRODUCT CODE	7513400
BATCH NUMBER	W025161
PRODUCTION DATE	11/2008
EXPIRY DATE	11/2010

TEST	METHOD	LIMITS	RESULTS
Appearance		Clear Colourless Liquid	clear colourless liquid
Density (20°C)	GM 0110 01	0.865 - 0.885	0.877

Date: 12/12/08

QA Manager: [Signature]

QA: E. Goughan

JD Approval: [Signature]

This document has been produced on behalf of JohnsonDiversey by authorised personnel and is valid unsigned.

CERTIFICATE OF IRRADIATION

Number: **B** 0391345 00

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013447
 Date Rec'd 27/11/08
 Date: 5/12/08

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B208120083	JD IPA BOTTLES 25-45kGy	527	B/No.W025161 ACTUAL....38.6 - 41.2....kGy DOSE
B...LE0010999 B208120083	SAMPLES	2	CUSTOMERS BATCH/CODE No.s AS ORDER No.P013447 DATED 26/11/08
Total:		529	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 12-Dec-08
Date Tested : 17-Dec-08
Date Completed : 31-Dec-08
Purchase Order No. : PO13517

Test Article : Clearklens IPA
Sample Code : Works Order: 25161
Batch Reference : 255
Laboratory Reference : LR-00814236 **Article Quantity** : 20 Bottles
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 306045
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.



W.A. Barker

Date 05 Jun 2009

K.A. Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00129848
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 12-Dec-08
Date Tested : 17-Dec-08
Date Completed : 17-Dec-08
Purchase Order No. : PO13517

Test Article : Clearklens IPA

Sample Code : Works Order: 25161

Batch Reference : 255


Laboratory Reference : LR-00814237

Article Quantity : 3 Bottles

Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2000 EU/mL
Dilution	*	MM110/01	1/40
Spiked Recoveries	*	MM110/02	102 Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL

This sample was tested at a 1/40 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.
Product Standard Data Sheet: FM02



WAB

Date 22 Dec 2008

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00129848
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited