



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS IPA VH01
PRODUCT CODE	7513400
BATCH NUMBER	WQ 25192
PRODUCTION DATE	01/2009
EXPIRY DATE	01/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless Liquid	clear colourless liquid
Density (20°C)	GM 0110 01	0.865 – 0.885	0.885

Date:.....02/02/09.....

QA.Manager:.....*[Signature]*.....

QA:.....*[Signature]*.....

JD Approval:.....*[Signature]*.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel and is valid unsigned.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013630
 Date Rec'd 22/01/09
 Date: 27/01/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B209010499	JD IPA BOTTLES 25-45kGy	497	B/No.W025192 ACTUAL....40.3 - 42.0....kGy DOSE
B2FLE0010999 B209010499	SAMPLES	2	CUSTOMERS BATCH/CODE No.5 AS ORDER No.P013630 DATED 21/1/09
Total		499	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drake*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 06-Feb-09
Date Tested : 12-Feb-09
Date Completed : 12-Feb-09
Purchase Order No. : 13693

Test Article : ClearKlens IPA Block 1
Sample Code : Batch No: 428
Batch Reference : WO25192
Laboratory Reference : LR-00827677 **Article Quantity** : 1
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2000 EU/mL
Dilution	*	MM110/01	1/40
Spiked Recoveries	*	MM110/02	107% Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM02

This sample was tested at a 1/ 50 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.



W.A. Barker

Date 19 Feb 2009

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN00131766
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS (INTERNAL USE ONLY)

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 06-Feb-09
Date Tested : 17-Feb-09
Date Completed : 03-Mar-09
Purchase Order No. : 13693

Test Article : ClearKlens IPA Block 1
Sample Code : Batch No: 428
Batch Reference : WO25192
Laboratory Reference : LR-00827676 **Article Quantity** : 20 x 900mL
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM03

Day Book Ref: 380023
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.



MB

Date 05 Mar 2009

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN00131766
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited