



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	W025197
PRODUCTION DATE	02/2009
EXPIRY DATE	02/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless Liquid	clear colourless liquid
Density (20°C)	GM 0110 01	0.876 – 0.896	0.885

Date: 05/03/09

QA Manager: [Signature]

QA: E. Gagnon

JD Approval: [Signature]

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013746
 Date Rec'd 19/02/09
 Date: 24/02/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B209020549	JD DE 900ML BOTTLES 25-45kGy	683	172 X B/No.W025362 - CB/No.FC470 340 X B/No.W025197 - CB/No.FC401 171 X B/No.W025751 - CB/No.FC470 ACTUAL....37.6 - 39.3....kGy DOSE
B2FLE0010999 B209020549	SAMPLES	7	CUSTOMERS BATCH/CODE No.s AS ORDER No.P013746 DATED 18/2/09
Total		690	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drace*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 06-Mar-09
Date Tested : 11-Mar-09
Date Completed : 13-Mar-09
Purchase Order No. : PO13783

Test Article : Clearklens DE Block 1
Sample Code :
Batch Reference : WO25197
Laboratory Reference : LR-00836266 **Article Quantity** : 1 x 900mLs
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.25 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	82% Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM01

This sample was tested at a 1/50 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.

Date 19 Mar 2009

GLP

Compliant
Laboratory

K.A. Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00132925
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 06-Mar-09
Date Tested : 10-Mar-09
Date Completed : 24-Mar-09
Purchase Order No. : PO13783

Test Article : Clearklens DE Block 1
Sample Code :
Batch Reference : WO25197
Laboratory Reference : LR-00836265 **Article Quantity** : 23 x 900mLs
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 380037
Volume Tested: 20x50mL
Test carried out according to current Pharmacopoeias.



WAB

Date 27 Mar 2009

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology