



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARLENs IPA VH01
PRODUCT CODE	7513400
BATCH NUMBER	W025337
PRODUCTION DATE	03/2009
EXPIRY DATE	03/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless Liquid	Clear colourless liquid
Density (20°C)	GM 0110 01	0.865 - 0.885	0.877

Date:..... 27/03/09

QA.Manager:..... 

QA:..... E. Galaym

JD Approval:..... 

This document has been produced on behalf of JohnsonDiversey by authorised personnel and is valid unsigned.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013790
 Date Rec'd 5/03/09
 Date: 12/03/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B209030240	JD IPA BOTTLES 25-45kGy	528	B/No.W025337 ACTUAL....39.8 - 41.9....kGy DOSE
B2FLE0010999 B209030240	SAMPLES	2	CUSTOMERS BATCH/CODE No.s AS ORDER No.P013790 DATED 4/3/09
Total		530	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drake*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 18-Mar-09
Date Tested : 27-Mar-09
Date Completed : 27-Mar-09
Purchase Order No. : 13838

Test Article : Clearlens IPA Block 1
Sample Code :
Batch Reference : WO25337
Laboratory Reference : LR-00839545 **Article Quantity** : 1
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2000 EU/mL
Dilution	*	MM110/01	1/40
Spiked Recoveries	*	MM110/02	Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM02

This sample was tested at a 1/40 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.



W.A.B.

Date 03 Apr 2009

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00133351
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS


Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 18-Mar-09
Date Tested : 20-Mar-09
Date Completed : 03-Apr-09
Purchase Order No. : 13838

Test Article : Clearklens IPA Block 1
Sample Code :
Batch Reference : WO25337
Laboratory Reference : LR-00839546 **Article Quantity** : 22
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 380041
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.



W.A.B.

Date 08 Apr 2009

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00133351
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited