



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS CLEANSINALD VH09 SC
PRODUCT CODE	7516429
BATCH NUMBER	W025514
PRODUCTION DATE	04 / 2009
EXPIRY DATE	04 / 2011

TEST	METHOD	LIMITS	RESULTS
Appearance (20°C)	Visual	Clear Yellow Solution	clear Yellow Liquid.
pH (1% soln)	GM 0130 01	12.0 – 13.0	12.1
Density (20°C)	GM 0110 01	1.04 – 1.06	1.058

Date:.....13/05/09.....

QA.Manager:.....*[Signature]*.....

QA:.....*[Signature]*.....

JD Approval:.....*[Signature]*.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013768
 Date Rec'd 26/02/09
 Date: 4/03/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010113 B209030065	D-BOTTLES/CAPS/BAGS 25-40kgv	4	2 X B/No.W025511 2 X B/No.N/A W 025512 / W025514.
B2FLE0010999 B209030065	SAMPLES	1	
Total		5	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Dace*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 15-Apr-09
Date Tested : 17-Apr-09
Date Completed : 01-May-09
Purchase Order No. : P013943

Test Article : Cleansinald SC Block 2
Sample Code : Works Order No. 25514
Batch Reference : Fluid Batch No. VLF99C01 08 056
Laboratory Reference : LR-00847346 **Article Quantity** : 20 Bottles
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM06

Day Book Ref: 484014
Volume Tested: All
Test carried out according to current Pharmacopoeias.

Suz Wood
Mrs S L Wood, BSc., CBiol., MBiol
Laboratory Manager - Pharmaceutical Microbiology

Date 12 May 2009



K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology