



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS TEGO 2000 SC VH25
PRODUCT CODE	7513387
BATCH NUMBER	W025576
PRODUCTION DATE	05 / 2009
EXPIRY DATE	05 / 2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Yellow Clear Liquid	Yellow clear liquid
pH (neat)	GM 0130 01	7.5 - 8.5	8.3
Density (20°C)	GM 0110 01	0.993 - 1.003	1.000

Date:.....15/05/09.....

QA.Manager:.....*[Signature]*.....

QA:.....*[Signature]*.....

JD Approval:.....*[Signature]*.....

*This document has been produced on behalf of JohnsonDiversey by authorised personnel.*

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster




**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P014009  
 Date Rec'd 30/04/09  
 Date: 7/05/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B209050071	DIVERSEY TEGO 2000 10x5x50 25-38kGy	98	B/No.W025576 - CB/No.ES68656878  ACTUAL <u>35.0 - 36.5</u> .....kGy DOSE
B2FLE0010999 B209050071	SAMPLES	5	  CUSTOMERS BATCH/CODE No.s AS ORDER No.P014009 DATED 29/4/09
B2FLE0010999 B209050071	SAMPLES	1	SFARES
Total		104	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED