






CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS CLEANSINALD SS VH09S
PRODUCT CODE	SKU. 7516432
BATCH NUMBER	W025582
PRODUCTION DATE	01/2009
EXPIRY DATE	01/2010

TEST	METHOD	LIMITS	RESULTS
Appearance (20°C)	Visual	Clear Solution	clear solution
pH (neat)	GM 0130 01	9.0 – 10.5	9.0
Density (20°C)	GM 0110 01	0.990 – 1.010	1.000

QA Manager:  Date: 22/01/09

QA:  Date: 22-01-09

JD Approval:  Date: 23/01/09

Date above relates to the date of signing of the document by each responsible person.

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013592
 Date Rec'd 12/01/09
 Date: 16/01/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010127 B209010300	C/SINALD SS -BLK-2-W025581 25-40kGy	3	B/No.W025581 -W025582 ACTUAL...28.5 - 37.5...kGy DOSE Validation Results CUSTOMERS BATCH/CODE No.s AS ORDER No.P013592 DATED 12/1/09
B2FLE001VAL	VALIDATION		SINGLE VALIDATION
Total		3	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C Drake*

For and on behalf of ISOTRON LIMITED


CERTIFICATE OF ANALYSIS

Mr J Jackson	Date Received	:	30-Jan-09
Flexible Medical Packaging Limited	Date Tested	:	04-Feb-09
Unit 8, Hightown	Date Completed	:	18-Feb-09
White Cross Industrial Estate	Purchase Order No.	:	13658
Lancaster LA1 4XS			

Test Article	: Cleansinald SS Block 2	Article Quantity	: 24 Bottles
Sample Code	: Works Order 25582		
Batch Reference	: 508		
Laboratory Reference	: LR-00825952		
Test Required	: Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.		

Test	*	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM06

Day Book Ref: 380020
 Volume Tested: 20 x 50mL
 Test carried out according to current Pharmacopoeias.



 Date 24 Feb 2009

K.A. Barker, C.Biol. M.I. Biol.,
 Business Manager - Microbiology

Sample Booking Number: CN\00131517

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited