

CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS TEGO 2000 SS VH25
PRODUCT CODE	SKU 7516427
BATCH NUMBER	W025651
PRODUCTION DATE	01/2009
EXPIRY DATE	01/2010 ✓

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Yellow Clear Liquid	clear liquid.
pH (neat)	GM 0130 01	7.5 - 8.5	7.7
Density (20°C)	GM 0110 01	0.990 - 1.010	1.000

Date:.....20.01.09.....

QA Manager: .....

QA:.....E. Gaughan.....

JD Approval: .....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS
 Cust. Ref: FQ13614
 Date Rec'd 20/01/09
 Date: 22/01/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
BZFLE0010130 B209010447	TEGO SS STERILE FILL 25-40kGy	9	3 X W025654 → W025655 (FINISHED GOODS) 3 X W025655 3 X W025656 MAX...36.1...MIN...28.8...kGy Light Vali MAX...35.6...MIN...26.9...kGy Heavy Vali
BZFLE001VAL	VALIDATION		SINGLE VALIDATION ON 7.6KG SINGLE VALIDATION ON 16.8KG
BZFLE001PRE	PREMIUM SERVICE		
Total		9	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *[Signature]*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 30-Jan-09
Date Tested : 06-Feb-09
Date Completed : 20-Feb-09
Purchase Order No. : 13658


Test Article : Tego 2000 SS Block 1
Sample Code : Works Order 25651
Batch Reference : 509
Laboratory Reference : LR-00825943 **Article Quantity** : 24 Bottles
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 380020
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.

Mrs S L Wood, BSc, CBiol., MIBiol
Laboratory Manager - Pharmaceutical Microbiology

Date 25 Feb 2009



K.A.Barker, C.Biol, M.J. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN00131514
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited