



**CERTIFICATE OF ANALYSIS**

PRODUCT NAME	CLEARKLENS TEGO 2000 SS VH25
PRODUCT CODE	SKU 7516428
BATCH NUMBER	W025652
PRODUCTION DATE	01/2009
EXPIRY DATE	01/2010 ✓

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Yellow Clear Liquid	clear liquid
pH (neat)	GM 0130 01	7.5 - 8.5	7.7
Density (20°C)	GM 0110 01	0.990 - 1.010	1.000

Date:.....26.10.09.....

QA Manager:..........

QA:.....E. Gough.....

JD Approval:..........

*This document has been produced on behalf of JohnsonDiversey by authorised personnel.*

FLE001

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS  
 Cust. Ref: PQ13614  
 Date Rec'd 20/01/09  
 Date: 22/01/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010130 B209010447	TEGO SS STERILE FILL 25-40kGy	9	3 X W025654 3 X W025655 - W025652 (FINISHED 46003) 3 X W025656  MAX...36.1...MIN...28.8...kGy Light Vali  MAX...35.6...MIN...26.9...kGy Heavy Vali
B2FLE001VAL	VALIDATION		SINGLE VALIDATION ON 7.6KG SINGLE VALIDATION ON 16.8KG
B2FLE001PRE	PREMIUM SERVICE		
Total		9	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *[Signature]*

For and on behalf of ISOTRON LIMITED



## CERTIFICATE OF ANALYSIS


Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

**Date Received** : 30-Jan-09  
**Date Tested** : 04-Feb-09  
**Date Completed** : 18-Feb-09  
**Purchase Order No.** : 13658

**Test Article** : Tego 2000  
**Sample Code** : Works Order 25652  
**Batch Reference** : 509  
**Laboratory Reference** : LR-00825963 **Article Quantity** : 24  
**Test Required** : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM06

Day Book Ref: 380020  
Volume Tested: 20 x 50mL  
Test carried out according to current Pharmacopoeias.



*WB*

Date 24 Feb 2009

K.A.Barker, C.Biol. M.I. Biol.,  
Business Manager - Microbiology

Sample Booking Number: CN\00131520

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

*This report shall not be reproduced except in full without the written approval of Wickham Laboratories Limited*