



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS TEGO 2000 SS VH25
PRODUCT CODE	SKU 7516428
BATCH NUMBER	W025653
PRODUCTION DATE	01/2009
EXPIRY DATE	01/2010

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Yellow Clear Liquid	clear liquid
pH (neat)	GM 0130 01	7.5 - 8.5	7.7
Density (20°C)	GM 0110 01	0.990 - 1.010	1.000

Date:.....26/10/09.....

QA Manager:..........

QA:.....E. Gargu.....

JD Approval:..........

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013614
 Date Rec'd 20/01/09
 Date: 22/01/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010130 B209010447	TEGO SS STERILE FILL 25-40kGy	9	3 X W025654 3 X W025655 3 X W025656 - W025653 (FINISHED GOODS) MAX...36.1...MIN...28.8...kGy Light Vali MAX...35.6...MIN...26.9...kGy Heavy Vali
B2FLE001VAL	VALIDATION		SINGLE VALIDATION ON 7.6KG SINGLE VALIDATION ON 16.8KG
B2FLE001PRE	PREMIUM SERVICE		
Total		9	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *[Signature]*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson	Date Received	:	30-Jan-09
Flexible Medical Packaging Limited	Date Tested	:	06-Feb-09
Unit 8, Hightown	Date Completed	:	20-Feb-09
White Cross Industrial Estate	Purchase Order No.	:	13658
Lancaster LA1 4XS			

Test Article : Tego 2000 SS Block 2
 Sample Code : Works Order 25653
 Batch Reference : 509
 Laboratory Reference : LR-00825961 **Article Quantity** : 24 Bottles
 Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 380020
 Volume Tested: 20 x 50mL
 Test carried out according to current Pharmacopoeias.



WAB

Date 26 Feb 2009

K.A.Barker, C.Biol. M.I. Biol.,
 Business Manager - Microbiology