



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS TEGO 2000 SC VH25
PRODUCT CODE	7513387
BATCH NUMBER	W025740
PRODUCTION DATE	✓ 02/2009
EXPIRY DATE	✓ 02/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Yellow Clear Liquid	Yellow clear liquid
pH (neat)	GM 0130 01	7.5 – 8.5	7.8
Density (20°C)	GM 0110 01	0.993 – 1.003	1.002

Date: 20.10.2009  
 QA Manager: [Signature]  
 QA: E. Gamm  
 JD Approval: [Signature]

*This document has been produced on behalf of JohnsonDiversey by authorised personnel.*

**CERTIFICATE OF IRRADIATION**    Number: **B**    0398542 00

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**  
 Roydsdale Way · Euoway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013730  
 Date Rec'd 12/02/09  
 Date: 23/02/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B209020373	DIVERSEY TEGO 2000 10x5x50 25-38kGy	99	20 X B/No.W025740 79 X B/No.W025194 CUSTOMER B/No.ES68757657  ACTUAL 32.8 - 33.1 .....kGy DOSE
B2FLE0010999 B209020373	SAMPLES	4	3 X B/No.W025740 1 X B/No.W025194  CUSTOMERS BATCH/CODE No.s AS ORDER No.P013730 DATED 11/2/09
B2FLE001MIN	MINIMUM CHARGE UPLIFT		
Total		103	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drace*

For and on behalf of ISOTRON LIMITED