



**CERTIFICATE OF ANALYSIS**

PRODUCT NAME	CLEARLENSS TEGO 2000 SS VH25
PRODUCT CODE	7516427
BATCH NUMBER	W025794 - W025791
PRODUCTION DATE	05/2009
EXPIRY DATE	05/2010

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Yellow Clear Liquid	clear liquid
pH (neat)	GM 0130 01	7.5 - 8.5	7.5
Density (20°C)	GM 0110 01	0.990 - 1.010	1.001

Date: 10/05/09.....  
 QA Manager: [Signature].....  
 QA: E. Gamm.....  
 JD Approval: [Signature].....

*This document has been produced on behalf of JohnsonDiversey by authorised personnel.*

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**  
 Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013971  
 Date Rec'd 23/04/09  
 Date: 27/04/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010130 B209040592	TEGO SS STERILE FILL 25-40kGy	16	12 X B/No. W025791 - W025794 4 X B/No. W025792  CUSTOMERS BATCH/CODE No.s AS ORDER No. P013971 DATED 22/4/09
Total		16	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drace*

For and on behalf of ISOTRON LIMITED



## CERTIFICATE OF ANALYSIS


Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

**Date Received** : 15-May-09  
**Date Tested** : 20-May-09  
**Date Completed** : 03-Jun-09  
**Purchase Order No.** : 14075

**Test Article** : Clearklens Tego 2000 SS VH25  
**Sample Code** : W025794 - W025791  
**Batch Reference** :  
**Laboratory Reference** : LR-00856048 **Article Quantity** : 24 x900mL  
**Test Required** : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 484038  
Volume Tested: 20 x 50mL  
Test carried out according to current Pharmacopoeias.



*MR*

Date 08 June 2009

K.A.Barker, C.Biol, M.I. Biol.,  
Business Manager - Microbiology

**Sample Booking Number:** CN\00135678

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*