



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS TEGO 2001 SS VH24
PRODUCT CODE	7513403
BATCH NUMBER	W025859
PRODUCTION DATE	03/2009
EXPIRY DATE	01/2010

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless to Pale Yellow Liquid	clear colourless liquid
Odour		Typical Slight	Typical
pH (neat)	GM 0130 01	8.5 – 9.2	8.8
Density (20°C)	GM 0110 01	0.998 – 1.001	1.001

Date:.....27/03/09.....

QA.Manager:.....*[Signature]*.....

QA:.....*E. Gavagan*.....

JD Approval:.....*[Signature]*.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: F013831
 Date Rec'd 17/03/09
 Date: 19/03/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010093 B209030502	TEGO 2001 SS 25-42kGy	55	B/No.W025859 - CUSTOMER B/No.283996 ACTUAL....38.4 - 39.6....kGy DOSE CUSTOMERS BATCH/CODE No.s AS ORDER No.P013831 DATED 16/3/09
B2FLE001MIN	MINIMUM CHARGE UPLIFT		
B2FLE001MIS	MISCELLANEOUS		PREMIUM RATE CHARGE
B2FLE0010999 B209030502	SAMPLES	2	
Total		57	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED