



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARLEN S TEGO 2001 SC VH24
PRODUCT CODE	7513402
BATCH NUMBER	W0 25861
PRODUCTION DATE	07/2009
EXPIRY DATE	07/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless to Yellow Liquid	clear colourless to yellow liquid
Odour		Typical	Typical
pH (neat)	GM 0130 01	8.5 – 9.5	8.6
Density (20°C)	GM 0110 01	1.060 – 1.070	1.066
Refractive Index (20°C)	GM 0120 01	1.374 – 1.384	1.380

Date:.....29/07/09.....
 QA.Manager:.....*[Signature]*.....
 QA:.....E. Gagnon.....29/07/09.....
 JD Approval:.....*[Signature]*.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P014293
 Date Rec'd 16/07/09
 Date: 23/07/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010091 B209070463	DIVERSEY TEG02001 10x5x100 25-42kGy	186	B/No.W025861 CUSTOMER B/No.286139 ACTUAL....32.7 - 33.5....kGy DOSE
B2FLE0010999 B209070463	SAMPLES	1	CUSTOMERS BATCH/CODE No.s AS ORDER No.P014293 DATED 16/7/09
Total		187	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED