



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS TEGO 2000 SC VH25
PRODUCT CODE	7513387
BATCH NUMBER	W025862
PRODUCTION DATE	06/2009
EXPIRY DATE	06/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	VISUAL	Yellow Clear Liquid	YELLOW CLEAR LIQUID
pH (neat)	GM 0130 01	7.5 - 8.5	7.6
Density (20°C)	GM 0110 01	0.993 - 1.003	0.999

Date: 25/06/09

QA Manager: [Signature]

QA: E. Gungun 25-06-09

JD Approval: [Signature] 25/06/09

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: 14168
Date Rec'd 11/06/09
Date: 18/06/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B209060329	DIVERSEY TEGO 2000 10x5x50 25-38kGy	98	B/No.W025862 - CUSTOMER B/No.287933 ACTUAL....33.4 - 33.9....kGy DOSE
B2FLE0010999 B209060329	SAMPLES	1	CUSTOMERS BATCH/CODE No.s AS ORDER No.P014168 DATED 11/6/09
Total		99	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *R Lygden*

For and on behalf of ISOTRON LIMITED