



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	W025878
PRODUCTION DATE	04 / 2009
EXPIRY DATE	04 / 2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless Liquid	clear colourless liquid
Density (20°C)	GM 0110 01	0.876 – 0.896	0.883

Date:.....24/04/09.....
 QA.Manager:.....*[Signature]*.....
 QA:.....*E. G...*.....
 JD Approval:.....*[Signature]*.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013950
 Date Rec'd 16/04/09
 Date: 22/04/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B209040454	JD IPA BOTTLES 25-45kGy	173	B/No.25878 - PRODUCT REF.7515783 ACTUAL....40.3 - 41.7....kGy DOSE
B2FLE0010999 B209040454	SAMPLES	2	CUSTOMERS BATCH/CODE No.s AS ORDER No.P013950 DATED 16/4/09
Total		175	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *[Signature]*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 28-Apr-09
Date Tested : 06-May-09
Date Completed : 06-May-09
Purchase Order No. : 13980-

Test Article : Clearklens DE Block 2
Sample Code : WO25878
Batch Reference : 551
Laboratory Reference : LR-00850710 **Article Quantity** : 1
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2500 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	82% Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM01

This sample was tested at a 1/50 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.

Mrs S L Wood BSc., C.Biol., M.Biol.
Laboratory Manager - Pharmaceutical Microbiology

Date 12 May 2009



K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 28-Apr-09
Date Tested : 06-May-09
Date Completed : 20-May-09
Purchase Order No. : 13980-


Test Article : Clearklens DE Block 2
Sample Code : WO25878
Batch Reference : 551
Laboratory Reference : LR-00850709 **Article Quantity** : 20
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM03

Day Book Ref: 484021
Volume Tested: 20x50mL
Test carried out according to current Pharmacopoeias.

Mrs S L Wood, BSc., CBiol., MIBiol
Laboratory Manager - Pharmaceutical Microbiology

Date 21 May 2009



K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00134875

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited