



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	W026273
PRODUCTION DATE	07/2009
EXPIRY DATE	07/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	VISUAL	Clear Colourless Liquid	Clear Colourless /19W1d
Density (20°C)	GM 0110 01	0.876 – 0.896	0.881

Date: 17/07/09
 QA Manager: [Signature]
 QA: E. Gamm 17.07.09
 JD Approval: [Signature]

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P014266
Date Rec'd 9/07/09
Date: 14/07/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B209070316	JD DE 900ML BOTTLES 25-45kGy	704	B/No.W026273 - CUSTOMER B/No.910 ACTUAL....36.1 - 37.4....kGy DOSE
B2FLE0010999 B209070316	SAMPLES	2	 CUSTOMERS BATCH/CODE No.s AS ORDER No.P014266 DATED 8/7/09
Total		706	Last Page 1 of 1

All in accordance with current
technical agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson

Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 20-Jul-09
Date Tested : 22-Jul-09
Date Completed : 22-Jul-09
Purchase Order No. : P014311

Test Article : Clearklens DE Block 1

Sample Code : 910

Batch Reference : WO26273

Laboratory Reference : LR-00873809

Article Quantity : 1 x 900mLs

Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.


Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2500 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	94% Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM01

This sample was tested at a 1/50 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.

Mr S L Wood FSc CBiol, MIBiol

Senior Analyst - Pharmaceutical Microbiology

Date 30 July 2009



K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00138398

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 20-Jul-09
Date Tested : 28-Jul-09
Date Completed : 11-Aug-09
Purchase Order No. : P014311

Test Article : Clearklens DE Block 1
Sample Code : 910
Batch Reference : WO26273
Laboratory Reference : LR-00873808 **Article Quantity** : 20 x 900mLs
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 574028
Volume Tested: 20x80mL
Test carried out according to current Pharmacopoeias.

GLP
Compliant
Laboratory

W.A.B.

Date 19 Aug 2009

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00138398
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited