



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	W026515
PRODUCTION DATE	09/2009
EXPIRY DATE	09/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless Liquid	clear colourless Liquid
Density (20°C)	GM 0110 01	0.876 – 0.896	0.884

Date:.....22/09/09.....

QA Manager:.....*[Signature]*.....

QA:.....*E. Gavanagh*.....

JD Approval:.....*[Signature]*.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P014522
Date Rec'd 11/09/09
Date: 15/09/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B209090295	JD DE 900ML BOTTLES 25-45kGy	1057	FMP BATCH NO:W026515-CUST NO:1132 ACTUAL....36.3 - 36.9....kGy DOSE CUSTOMER BATCH/CODE NO AS ORDER NO P014522 DATED 09.09.09
B2FLE0010999 B209090295	SAMPLES	2	
Total		1059	Last Page 1 of 1

All in accordance with current
technical agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature:

J. Gaffey

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson

Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 22-Sep-09
Date Tested : 30-Sep-09
Date Completed : 30-Sep-09
Purchase Order No. : -

Test Article : Clearklens DE
Sample Code : Works Order 26515
Batch Reference : 1132
Laboratory Reference : LR-00890523 Article Quantity : 1
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.


Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2500 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM01

This sample was tested at a 1/50 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.

Mr K.A. Barker C.Biol. M.I. Biol.,
Business Manager

Date 07 Oct 2009

Mrs S.L. Wood, B.Sc. (Hons), C.Biol., M.I. Biol.,
Laboratory Manager - Pharmaceutical Microbiology



Sample Booking Number: CN00140933
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 22-Sep-09
Date Tested : 01-Oct-09
Date Completed : 15-Oct-09
Purchase Order No. : 14581

Test Article : Clearlens
Sample Code : Works Order 26515
Batch Reference : 1132
Laboratory Reference : LR-00890522 Article Quantity : 20
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 649016

Volume Tested: 20 x 50mL

Test carried out according to current Pharmacopoeias.

GLP

Compliant
Laboratory

SL Wood
SL Wood, BSc, C.Biol, MSB
Lab Manager, Pharm Micro

Date 15 Oct 2009

PP K.A. Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00140933

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited