



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	NO 26750
PRODUCTION DATE	10 / 2009
EXPIRY DATE	10 / 2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless Liquid	Clear Colourless Liquid
Density (20°C)	GM 0110 01	0.876 – 0.896	0.885

Date: 19. 10. 2009

QA.Manager: [Signature]

QA: E. Gwynn

JD Approval: [Signature]

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

CERTIFICATE OF IRRADIATION

Number: **B** 0423510 00

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS
 P014640

Cust. Ref: P014640
 Date Rec'd 6/10/09
 Date: 9/10/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B209100159	JD DE 900ML BOTTLES 25-45kGy	712	B/No.W026750 - CUSTOMER B/No.1273 ACTUAL....36.1....37.3....kGy DOSE
B2FLE0010999 B209100159	SAMPLES	2	CUSTOMERS BATCH/CODE No.s AS ORDER No.P014640 DATED 5/10/09
Total		714	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Goffey*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 20-Oct-09
Date Tested : 25-Oct-09
Date Completed : 06-Nov-09
Purchase Order No. : 14697

Test Article : Clearklens DE Block 1
Sample Code : WO26750
Batch Reference : 1273
Laboratory Reference : LR-00898629 **Article Quantity** : 20 Bottles
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03


Day Book Ref: 649033
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.

Suz Wood

Mrs S.L. Wood, BSc., CBiol., MSB

Laboratory Manager - Microbiology

Date 12 Nov 2009



K.A. Barker, C.Biol, MSB
Business Manager - Microbiology

Sample Booking Number: CN00142117
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 20-Oct-09
Date Tested : 27-Oct-09
Date Completed : 27-Oct-09
Purchase Order No. : 14697

Test Article : Clearlens DE Block 1
Sample Code : WO26750
Batch Reference :
Laboratory Reference : LR-00898630 **Article Quantity** : 1 Bottle
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2500 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM01

This sample was tested at a 1/50 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.

Mr K A Barker, CBiol., MSB
Business Manager

WAB

Date 29 Oct 2009



Mrs S.L. Wood, B.Sc. (Hons), C.Biol., MSB.,
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN\00142117

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited