




**CERTIFICATE OF ANALYSIS**


<b>PRODUCT NAME</b>	<b>CLEARKLENS TEGO 2000 SC VH25</b>
<b>PRODUCT CODE</b>	<b>7513387</b>
<b>BATCH NUMBER</b>	W026753
<b>PRODUCTION DATE</b>	10/2009
<b>EXPIRY DATE</b>	10/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Yellow Clear Liquid	yellow clear liquid
pH (neat)	GM 0130 01	7.5 - 8.5	7.5
Density (20°C)	GM 0110 01	0.993 - 1.003	1.003

Date: 09/11/09.....

QA Manager: .....

QA: .....

JD Approval: .....

*This document has been produced on behalf of JohnsonDiversey by authorised personnel.*

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P014677  
 Date Rec'd 15/10/09  
 Date: 26/10/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B209100444	DIVERSEY TEGO 2000SC 50ml 25-38kGy	100	B/No.W026753 - CUSTOMER B/No.309113  ACTUAL DOSE 35.2 kGy  CUSTOMER'S BATCH/CODE No.5 AS ORDER No.P014677 DATED 14/10/09
B2FLE0010999 B209100444	SAMPLES	1	CLEARKLENS PLUS  CERTIFIED TO THE OUTSIDE OF THE BOX ONLY: 37.5 kGy.
B2FLE001MIN	MINIMUM CHARGE UPLIFT		
B2FLE001VAL	VALIDATION		SINGLE VALIDATION ON TEGO 2000
Total		101	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED