



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS TEGO 2000 SC VH25
PRODUCT CODE	7513387
BATCH NUMBER	W0 26828
PRODUCTION DATE	11/2009
EXPIRY DATE	11/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	VISUAL	Yellow Clear Liquid	yellow clear liquid
pH (neat)	GM 0130 01	7.5 – 8.5	7.9
Density (20°C)	GM 0110 01	0.993 – 1.003	1.003

Date:..... 27/11/09

QA.Manager: *[Signature]*

QA:..... *E. Gausman*

JD Approval: *[Signature]*

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited
 Roydsdale Way · Euoway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P014824
 Date Rec'd 19/11/09
 Date: 23/11/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B209110376	DIVERSEY TEGO 2000SC 50ml 25-42kGy	101	B/No.W026828 CUSTOMER B/No.303485 ACTUAL DOSE 33.8-34.0 kGy Processed under Vali Ref 4.0716
B2FLE0010999 B209110376	SAMPLES	2	 CUSTOMERS BATCH/CODE No.s AS ORDER No.P014824 DATED 18/11/09
Total		103	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED