



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	NO 26853
PRODUCTION DATE	2009 / 09
EXPIRY DATE	2011 / 09

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless Liquid	clear Colourless Liquid
Density (20°C)	GM 0110 01	0.876 - 0.896	0.882

Date:.....12/10/09.....

QA.Manager:.....*[Signature]*.....

QA:.....*E. Gagnier*.....

JD Approval:.....*[Signature]*.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P014623
 Date Rec'd 1/10/09
 Date: 6/10/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B209100070	JD IPA BOTTLES 25-45kGy	179	B/No.W026853 ACTUAL.....38.3.....39.3.....kGy DOSE
B2FLE0010999 B209100070	SAMPLES	2	CUSTOMERS BATCH/CODE No.s AS ORDER No.P014623 DATED 30/9/2009
Total		181	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 08-Oct-09
Date Tested : 12-Oct-09
Date Completed : 12-Oct-09
Purchase Order No. : PO14653


Test Article : Clearklens IPA Block 1
Sample Code :
Batch Reference : WO26853
Laboratory Reference : LR-00895004 **Article Quantity** : 1 x 900mL
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2000 EU/mL
Dilution	*	MM110/01	1/40
Spiked Recoveries	*	MM110/02	Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM02

This sample was tested at a 1/40 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.

Date

20 Oct 2009



K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00141616
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 08-Oct-09
Date Tested : 21-Oct-09
Date Completed : 04-Nov-09
Purchase Order No. : PO14653

Test Article : Clearklens IPA Block 1
Sample Code :
Batch Reference : WO26853
Laboratory Reference : LR-00895005 **Article Quantity** : 20 x 900mL
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.


Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM03

Day Book Ref: 649026
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.

SL Wood

Mrs SL Wood, BSc., CBiol., MSB
Laboratory Manager - Pharmaceutical Microbiology

Date 06 Nov 2009



K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN00141616
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited