

Remove the cap with filter
Base last off with filter


JohnsonDiversey
Clean is just the beginning





CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS BI-SPORE BASE
PRODUCT CODE	7514852
BATCH NUMBER	W0 26857
DATE OF MANUFACTURE	10/2009
EXPIRY DATE	10/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	1 (a)	Pale Blue Liquid	Pale Blue Liquid
pH (neat)	G14 (002)	1.5 – 3.5	1.9
Density (20°C)	G12 (001)	1.020 – 1.060	1.030

QA.Manager:  Date: 29/10/09

QA.:  Date: 29-10-09

JD Approval:  Date: 29/10/09

Date above relates to the date of signing of the document by each responsible person.

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

Application must not with filter
Base last off with filter

JohnsonDiversey
Clean is just the beginning



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS BI-SPORE ACTIVATOR
PRODUCT CODE	7514852
BATCH NUMBER	W0 26857
DATE OF MANUFACTURE	10/2009
EXPIRY DATE	10/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	1 (a)	Clear Colourless Liquid	clear colourless liquid
pH (neat)	G14 (002)	11.5 – 13.0	11.5
Density (20°C)	G12 (001)	1.015 – 1.035	1.016

QA.Manager: [Signature] Date: 29/10/09

QA: [Signature] Date: 29.10.09

JD Approval: [Signature] Date: 29/10/09

Date above relates to the date of signing of the document by each responsible person.

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

Base lost off cloth filter

JohnsonDiversey
Clean is just the beginning



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS BI-SPORE WORKING SOLUTION
PRODUCT CODE	7514852
BATCH NUMBER	WO 26857
DATE OF MANUFACTURE	10/2009
EXPIRY	24 hours from generation of solution

TEST	METHOD	LIMITS	RESULTS
Appearance	1 (a)	Clear Yellow Liquid	clear yellow liquid
pH (neat)	G14 (002)	2.5 – 5.5	3
Density (20°C)	G12 (001)	1.000 – 1.010	1.001
Chlorine Dioxide Content		230 – 280ppm	236.25ppm

QA.Manager: *[Signature]* Date: 29/10/09

QA: *[Signature]* Date: 29.10.09

JD Approval: *[Signature]* Date: 29/10/09

Date above relates to the date of signing of the document by each responsible person.

This document has been produced on behalf of JohnsonDiversey by authorised personnel.



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 26-Oct-09
Date Tested : 30-Oct-09
Date Completed : 13-Nov-09
Purchase Order No. : PO14719

Test Article : ClearKlens Bi Spore Activator
Sample Code : Work Order No. 26857
Batch Reference : 92870
Laboratory Reference : LR-00900336 **Article Quantity** : 40 x 100mL
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM04

Day Book Ref: 649035
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.

Mr K A Barker, CBiol., MSB
Business Manager

MKB

Date 20 Nov 2009



Mrs S.L. Wood, B.Sc. (Hons), C.Biol., MSB.,
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN00142361

Certificate of Analysis - OSMM

Page 1 of 2

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 26-Oct-09
Date Tested : 30-Oct-09
Date Completed : 13-Nov-09
Purchase Order No. : PO14719

Test Article : ClearKlens Bi Spore Base
Sample Code : Work Order No. 26857
Batch Reference : 92870
Laboratory Reference : LR-00900337 **Article Quantity** : 40 x 100mL
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.


Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM05

Day Book Ref: 649035
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.

Mr K A Barker, CBiol., MSB
Business Manager

WAB

Date 20 Nov 2009



Mrs S.L. Wood, B.Sc. (Hons), C.Biol., MSB.,
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN\00142361
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 2 of 2

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited