



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARLENs TEGO 2001 SC VH24
PRODUCT CODE	7513402
BATCH NUMBER	W026934
PRODUCTION DATE	10/2009
EXPIRY DATE	10/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	VISUAL	Clear Colourless to Yellow Liquid	clear colourless to yellow liquid
Odour		Typical	TYPICAL
pH (neat)	GM 0130 01	8.5 - 9.5	9.0
Density (20°C)	GM 0110 01	1.060 - 1.070	1.068
Refractive Index (20°C)	GM 0120 01	1.374 - 1.384	1.383

Date: 15/11/09

QA Manager: [Signature]

QA: E Gausin

JD Approval: [Signature]

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P014743
 Date Rec'd 29/10/09
 Date: 6/11/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010091 B209110106	DIVERSEY TEG02001 10x5x100 25-42kGy	191	B/No. W026934-CUSTOMER B/No.309187 ACTUAL DOSE 30.8 - 32.6 kGy
B2FLE0010999 B209110106	SAMPLES	2	
B2FLE0010091 B209110106	DIVERSEY TEG02001 10x5x100 25-42kGy	9	B/No. W027086-CUSTOMER B/No.309185 ACTUAL DOSE 30.8 - 32.6 kGy
B2FLE0010999 B209110106	SAMPLES	2	
B2FLE001VAL	VALIDATION		SINGLE VALIDATION (4.0706) ON B/No. W027086
Total		204	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED