


CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	WO 27122
PRODUCTION DATE	01/2010
EXPIRY DATE	01/2012

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless Liquid	Clear colourless liquid
Density (20°C)	GM 0110 01	0.876 - 0.896	0.888

Date: 21/01/2010

QA Manager: 

QA: 

JD Approval: 

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited
Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P014961
Date Rec'd 14/01/10
Date: 18/01/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B210010291	JD DE 900ML BOTTLES 25-45kGy	693	346 X B/No.W027122 CUSTOMER B/No1272 347 X B/No.W027471 CUSTOMER B/No1334 ACTUAL DOSE....36.1 - 37.7....kGy
B2FLE0010999 B210010291	SAMPLES	12	6 X B/No.W027122 CUSTOMER B/No.1272 6 X B/No.W027471 CUSTOMER B/No.1334 CUSTOMERS BATCH/CODE No.s AS ORDER No.P014961 DATED 13/1/10
Despatch Text: 14PLTS			
Total		705	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Gaffney*
For and on behalf of ISOTRON LIMITED

CERTIFICATE OF ANALYSIS

 Mr J Jackson
 Flexible Medical Packaging Limited
 Unit 8, Hightown
 White Cross Industrial Estate
 Lancaster LA1 4XS


Date Received : 21-Jan-10
Date Tested : 25-Jan-10
Date Completed : 25-Jan-10
Purchase Order No. : 15002

Test Article : Clearklens DE Block 1
Sample Code : W027122
Batch Reference : 1272
Laboratory Reference : LR-00924282 **Article Quantity** : 1 x 900mL
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test	Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	* MM110	<0.2500 EU/mL
Dilution	* MM110/01	1/50
Spiked Recoveries	* MM110/02	63% Valid
Endotoxin Spike Level	* MM110/09	0.5 EU/mL
Product Standard Data Sheet	* MQ002/01	FM01

 This sample was tested at a 1/ 50 dilution and spiked with endotoxin to check for interference.
 Test carried out according to current Pharmacopoeias.



 Date 27 Jan 2010

 K.A. Barker, C.Biol, M.I. Biol.,
 Business Manager - Microbiology

Sample Booking Number: CN\00145682
 Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited


CERTIFICATE OF ANALYSIS

 Mr J Jackson
 Flexible Medical Packaging Limited
 Unit 8, Hightown
 White Cross Industrial Estate
 Lancaster LA1 4XS

Date Received : 21-Jan-10
Date Tested : 28-Jan-10
Date Completed : 11-Feb-10
Purchase Order No. : 15002

Test Article : Clearklens DE Block I
Sample Code : W027122
Batch Reference : 1272
Laboratory Reference : LR-00924281 **Article Quantity** : 20 x 900mL
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM03

 Day Book Ref: 745034
 Volume Tested: 20x50mL
 Test carried out according to current Pharmacopoeias.



 Date 12 Feb 2010

 K.A.Barker, C.Biol, M.I. Biol.,
 Business Manager - Microbiology

Sample Booking Number: CN00145682

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited