



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	W027471
PRODUCTION DATE	01/2010
EXPIRY DATE	01/2012

TEST	METHOD	LIMITS	RESULTS
Appearance	VISUAL	Clear Colourless Liquid	clear colourless liquid
Density (20°C)	GM 0110 01	0.876 - 0.896	0.887

Date:.....21/01/2010.....
 QA.Manager:.....*[Signature]*.....
 QA:.....E. Garayin.....
 JD Approval:.....*[Signature]*.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P014961
 Date Rec'd 14/01/10
 Date: 18/01/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B210010291	JD DE 900ML BOTTLES 25-45kGy	693	346 X B/No.W027122 CUSTOMER B/No1272 347 X B/No.W027471 CUSTOMER B/No1334 ACTUAL DOSE....36.1 - 37.7....kGy
B2FLE0010999 B210010291	SAMPLES	12	6 X B/No.W027122 CUSTOMER B/No.1272 6 X B/No.W027471 CUSTOMER B/No.1334 CUSTOMERS BATCH/CODE No.s AS ORDER No.P014961 DATED 13/1/10
Despatch Text: 14PLTS			
Total		705	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Giffey*
 For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS


Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 21-Jan-10
Date Tested : 25-Jan-10
Date Completed : 25-Jan-10
Purchase Order No. : 15003

Test Article : Clearklens DE Block 1
Sample Code : W027471
Batch Reference : 1334
Laboratory Reference : LR-00924278 Article Quantity : 1 x 900mL
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2500 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	93% Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM01

This sample was tested at a 1/ dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.



Date 27 Jan 2010

K.A. Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00145679
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 21-Jan-10
Date Tested : 27-Jan-10
Date Completed : 10-Feb-10
Purchase Order No. : 15003

Test Article : Clearklens DE Block 1
Sample Code : W027471
Batch Reference : 1334
Laboratory Reference : LR-00924277 **Article Quantity** : 20 x 900mL
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM03

Day Book Ref: 745034
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.

WAS

Date 12 Feb 2010



K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00145679
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited