

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P015166  
 Date Rec'd 4/03/10  
 Date: 9/03/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B210030106	DIVERSEY TEGO 2000SC 50ml 25-42kGy	96	48 X B/No.W027560 48 X B/No.W027559 CUSTOMER B/No.324762  ACTUAL DOSE....33.3 - 33.6....kGy
B2FLE0010999 B210030106	SAMPLES	4	CUSTOMER BATCH/CODE No.s AS ORDER No P015166 DATED 3/3/10
B2FLE001MIN	MINIMUM CHARGE UPLIFT		
	Despatch Text: 2PLTS		
	Total	100	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Laffey*

For and on behalf of ISOTRON LIMITED




**CERTIFICATE OF ANALYSIS**

<b>PRODUCT NAME</b>	CLEARKLENS TEGO 2000 SC VH25
<b>PRODUCT CODE</b>	7514341
<b>BATCH NUMBER</b>	W027560
<b>PRODUCTION DATE</b>	03/2010
<b>EXPIRY DATE</b>	03/2012

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Yellow Clear Liquid	yellow clear liquid
pH (neat)	GM 0130 01	7.5 - 8.5	7.5
Density (20°C)	GM 0110 01	0.993 - 1.003	1.003

Date: <sup>(24)</sup> ~~13/02~~ 12/03/2010

QA Manager: 

QA: E. Gaurav

JD Approval: 

*This document has been produced on behalf of JohnsonDiversey by authorised personnel.*