



Contact address:

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 7/4/10

Product Name	CLEARKLENS DE VH29
Product Code	6069643
Batch Number	W027561
Production Date	03/2010
Expiry Date	03/2012

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Liquid	clear liquid
Specific Gravity (20°C)	JDM004	0.880 - 0.888	0.887

JOHN JACKSON

Signature / name
On behalf of Manufacturing Location



FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited
Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P015226
Date Rec'd 18/03/10
Date: 23/03/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B210030427	JD DE 900ML BOTTLES 25-45kGy	378	B/No.W027561 - CUSTOMER B/No.1598 ACTUAL DOSE....36.0 - 37.4....kGy
B2FLE0010999 B210030427	SAMPLES	5	 CUSTOMERS BATCH/CODE No.s AS ORDER No.P015226 DATED 17/3/04
Despatch Text: 7PLTS			
Total		383	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Gaffey*

For and on behalf of ISOTRON LIMITED

CERTIFICATE OF IRRADIATION

Number: **B** 0442018 00

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euoway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P015249
Date Rec'd 25/03/10
Date: 29/03/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B210030542	JD DE 900ML BOTTLES 25-45kGy	320	B/No.W027561 - CUSTOMER B/No.1598 ACTUAL DOSE....35.9 - 38.2....kGy
B2FLE0010999 B210030542	SAMPLES	1	CUSTOMERS BATCH/CODE No.s AS ORDER No.P015249 DATED 24/3/10
Despatch Text: 6PLTS			
Total		321	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 29-Mar-10
Date Tested : 06-Apr-10
Date Completed : 06-Apr-10
Purchase Order No. : PO15254

Test Article : ClearKlens JD DE Block 1
Sample Code : Work Order No. 27561
Batch Reference : 1598
Laboratory Reference : LR-00945488 **Article Quantity** : 1Bottle+3 Spare
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2500 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	Valid 98%
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM01

Test carried out according to current Pharmacopoeias.

Mr H.A. Burke, CBiol, MSB
Business Manager

Date 13 April 2010



Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN\00148496
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS


Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 29-Mar-10
Date Tested : 07-Apr-10
Date Completed : 21-Apr-10
Purchase Order No. : PO15254

Test Article : ClearKlens JD DE Block 1
Sample Code : Work Order No. 27561
Batch Reference : 1598
Laboratory Reference : LR-00945487 **Article Quantity** : 20 Bottles
Test Required : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 800020
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.



Suz Wood

Date 22 April 2010

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN\00148496

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited