

Base test off with filter

**JohnsonDiversey**  
Clean is just the beginning



Contact address:

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

### CERTIFICATE OF ANALYSIS

Date: 19.04.10

<b>Product Name</b>	CLEARKLENS BI-SPORE BASE SOLUTION VH26
<b>Product Code</b>	7514852
<b>Batch Number</b>	WO 27562
<b>Production Date</b>	03/2010
<b>Expiry Date</b>	03/2012

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Blue Liquid	clear blue liquid
pH (100% 20°C)	JDM001	1.5 - 3.5	1.9
Specific Gravity (20°C)	JDM004	1.020 - 1.060	1.030

*Gammisho*

Signature / name  
On behalf of Manufacturing Location



17/11/10  
Base lost off with filter



Contact address:

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

### CERTIFICATE OF ANALYSIS

Date: 19.04.10

<b>Product Name</b>	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26
<b>Product Code</b>	7514852
<b>Batch Number</b>	WO 27562
<b>Production Date</b>	03/2010
<b>Expiry Date</b>	03/2012

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Colourless Liquid	clear colourless liquid
pH (100% 20°C)	JDM001	11.5 - 13.5	11.5
Specific Gravity (20°C)	JDM004	1.015 - 1.035	1.016

*Gueni esto*

Signature / name  
On behalf of Manufacturing Location



ACTIVITY: 1001 11 1000 1110  
Base last off with filter



Contact address:  
JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

### CERTIFICATE OF ANALYSIS

Date: 19.04.10

<b>Product Name</b>	CLEARKLENS BI-SPORE COMBINED SOLUTION VH26
<b>Product Code</b>	7514852
<b>Batch Number</b>	W0 27562
<b>Production Date</b>	03/2010
<b>Expiry Date</b>	03/2012

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Yellow Liquid	clear yellow liquid
pH (100% 20°C)	JDM001	2.5 - 5.5	2.8
Specific Gravity (20°C)	JDM004	1.000 - 1.010	1.001
Potential Chlorine Dioxide	Titration	230 - 280ppm	256.5ppm

*Steven 'isho*

Signature / name  
On behalf of Manufacturing Location



FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**  
 Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS  
 Cust. Ref: P015127  
 Date Rec'd 18/02/10  
 Date: 26/02/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010113 B210020442	JD-BOTTLES/CAPS/BAGS 25-40kGy	12	B/No.W027558  CUSTOMERS BATCH/CODE No.s AS ORDER No.P015127 DATED 17/2/10
Despatch Text:  2PLTS		Total	12    Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Giffen*  
 For and on behalf of ISOTRON LIMITED



**CERTIFICATE OF ANALYSIS**

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

**Date Received** : 12-Mar-10  
**Date Tested** : 24-Mar-10  
**Date Completed** : 07-Apr-10  
**Purchase Order No.** : PO15208


**Test Article** : Bi Spore Base  
**Sample Code** : Works Order No: 27562  
**Batch Reference** : 00540  
**Laboratory Reference** : LR-00940323 **Article Quantity** : 20  
**Test Required** : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/2
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/2
Product Standard Data Sheet	*	MQ002/01	FM05

Day Book Ref: 800013  
Volume Tested: All  
Test carried out according to current Pharmacopoeias.

Mr K A Barker, CBiol., MSB  
Business Manager

Date 14 Apr 2010



Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Laboratory Manager - Pharmaceutical Microbiology

**Sample Booking Number:** CN\00147842  
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 2

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*



**CERTIFICATE OF ANALYSIS**

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

**Date Received** : 12-Mar-10  
**Date Tested** : 24-Mar-10  
**Date Completed** : 07-Apr-10  
**Purchase Order No.** : PO15208

**Test Article** : Bi Spore Activator  
**Sample Code** : Works Order No: 27562  
**Batch Reference** : 00541  
**Laboratory Reference** : LR-00940324 **Article Quantity** : 20  
**Test Required** : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.


Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/2
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/2
Product Standard Data Sheet	* MQ002/01	FM04

Day Book Ref: 800013  
Volume Tested: All  
Test carried out according to current Pharmacopoeias.

Mr K A Barker, CBiol., MSB  
Business Manager

*WAB*

Date 14 Apr 2010



Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Laboratory Manager - Pharmaceutical Microbiology

**Sample Booking Number:** CN\00147842  
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 2 of 2

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*