

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P015127
 Date Rec'd 18/02/10
 Date: 23/02/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010093 B210020436	TEGO 2001 SS 25-42kGy	62	B/No.W027606 CUSTOMER B/No.7515138 ACTUAL DOSE...37.9 - 38.2...kGy
B2FLE0010999 B210020436	SAMPLES	6	 CUSTOMERS BATCH/CODE No.s AS ORDER No.P015127 DATED 17/2/10
Despatch Text: 2PLTS			
Total		68	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Giffey*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS TEGO 2001 SS VH24 (DEV00155)
PRODUCT CODE	7513403
BATCH NUMBER	W027606
PRODUCTION DATE	02/2010
EXPIRY DATE	12/2010

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless to Pale Yellow Liquid	Clear Colourless to Pale yellow liquid
Odour	Typical	Typical Slight	Typical
pH (neat)	GM 0130 01	8.5 - 9.2	9.1
Density (20°C)	GM 0110 01	0.998 - 1.001	1.001

Date: 03/03/2010

QA Manager: [Signature]

QA: [Signature]

JD Approval: [Signature]

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

CERTIFICATE OF ANALYSIS


Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 04-Mar-10
Date Tested : 09-Mar-10
Date Completed : 23-Mar-10
Purchase Order No. : PO15761

Test Article : Clearklens Tego 2001 SS Block 1
Sample Code : Works Order No 27606
Batch Reference : 7515738
Laboratory Reference : LR-00937689 **Article Quantity** : 20 bottles
Test Required : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM03

Day Book Ref: 800008
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.



Suz Wood

Date 25 March 2010

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN\00147495
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited