

Contact address:

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

**CERTIFICATE OF ANALYSIS**

Date: 10/5/10

<b>Product Name</b>	CLEARKLENS DE VH29
<b>Product Code</b>	6069643
<b>Batch Number</b>	WG 27856
<b>Production Date</b>	04/2010
<b>Expiry Date</b>	04/2012

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Liquid	Clear liquid
Specific Gravity (20°C)	JDM004	0.880 - 0.888	0.880

Signature / name *C Pascoe*  
On behalf of Manufacturing Location





FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**  
 Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P015369  
 Date Rec'd 29/04/10  
 Date: 4/05/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B210050018	JD DE 900ML BOTTLES 25-45kGy	486	B/No.W027856 - CUSTOMER B/No.1772  ACTUAL DOSE....36.1 - 37.2....kGy
B2FLE0010999 B210050018	SAMPLES	5	CUSTOMERS BATCH/CODE No.s AS ORDER No.P015369 DATED 28/4/10
	Despatch Text:  9PLTS		
	Total	491	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Offley*

For and on behalf of ISOTRON LIMITED



**CERTIFICATE OF ANALYSIS**

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

**Date Received** : 10-May-10  
**Date Tested** : 12-May-10  
**Date Completed** : 12-May-10  
**Purchase Order No.** : 15398

**Test Article** : ClearKlens DE Block 1  
**Sample Code** :  
**Batch Reference** : W027856  
**Laboratory Reference** : LR-00959763 **Article Quantity** : 1 x 900mL  
**Test Required** : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.25 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	Valid 117%
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM01

Test carried out according to current Pharmacopoeias.



*Suz Wood*

Date 14 May 2010

Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Laboratory Manager - Pharmaceutical Microbiology



## CERTIFICATE OF ANALYSIS


Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

Date Received : 10-May-10  
Date Tested : 17-May-10  
Date Completed : 31-May-10  
Purchase Order No. : 15398

Test Article : ClearKlens DE Block 1  
Sample Code :  
Batch Reference : WO27856  
Laboratory Reference : LR-00959762 Article Quantity : 20 x 900mL  
Test Required : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM03

Day Book Ref: 800040  
Volume Tested: 20x50mL  
Test carried out according to current Pharmacopoeias.



*Suz Wood*

Date 10 June 2010

Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN\00150277  
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*