



Contact address:

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 27.05.2010

Product Name	CLEARKLENS IPA VH1
Product Code	7514372
Batch Number	No 27857
Production Date	05/2010
Expiry Date	05/2012

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Liquid	Clear
Specific Gravity (20°C)	JDM004	0.872 - 0.883	0.878

Signature / name *Ampelike Partyishe*
On behalf of Manufacturing Location



FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P015454
 Date Rec'd 20/05/10
 Date: 26/05/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B210050508	JD IPA BOTTLES 25-45kGy	706	351 X B/No.W027857 - CB/No.1701 355 X B/No.W027811 - CB/No.1558 ACTUAL DOSE....37.5 - 38.4....kGy
B2FLE0010999 B210050508	SAMPLES	11	5 X B/No.W027857 6 X B/No.W027811 CUSTOMERS BATCH/CODE No.s AS ORDER No.P015454 DATED 20/5/10
Despatch Text: 14PLTS			
Total		717	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS


Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 01-Jun-10
Date Tested : 07-Jun-10
Date Completed : 07-Jun-10
Purchase Order No. : 15479

Test Article : ClearKlens IPA Block 2
Sample Code :
Batch Reference : W027857
Laboratory Reference : LR-00967139 **Article Quantity** : 1x 900mL
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2000 EU/mL
Dilution	*	MM110/01	1/40
Spiked Recoveries	*	MM110/02	Valid 84%
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM02

Test carried out according to current Pharmacopoeias.



Suzanne Wood

Date *10 June 2010*

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN00151187
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 01-Jun-10
Date Tested : 10-Jun-10
Date Completed : 24-Jun-10
Purchase Order No. : 15479

Test Article : ClearKlens IPA Block 2
Sample Code :
Batch Reference : WO27857
Laboratory Reference : LR-00967138 **Article Quantity** : 20 x 900mL
Test Required : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM03

Day Book Ref: 900006
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.



Suz Wood

Date 25 June 2010

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN\00151187
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited