

Contact address:

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

## CERTIFICATE OF ANALYSIS

Date: 24/6/10

<b>Product Name</b>	CLEARKLENS TEGO 2000SS VH25S
<b>Product Code</b>	7516428
<b>Batch Number</b>	W028159
<b>Production Date</b>	06/2010
<b>Expiry Date</b>	06/2011

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Liquid	clear liquid
pH 100% (20°C)	JMD001	7.3 - 8.2	7.3
Specific Gravity (20°C)	JMD004	0.990 - 1.010	0.998

Signature / name *C Pascoe*  
On behalf of Manufacturing Location



FLE001

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster

Page 1



**Isotron Limited**

Roydsdale Way · Euoway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P015495  
 Date Rec'd 10/06/10  
 Date: 14/06/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010130 B210060367	TEGO SS STERILE FILL 25-40kGy         Despatch Text:  2PLTS       <p style="text-align: right;">Total</p>	<p style="text-align: center;">6</p>	2 X B/No.W028155 4 X B/No.W028154  CUSTOMERS BATCH/CODE No.s AS ORDER No.P015495 DATED 9/6/10  ACTUAL DOSE....36.0 - 37.7....kGy         <p style="text-align: right;">Last Page 1 of 1</p>

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *T. Giffen*

For and on behalf of ISOTRON LIMITED



## CERTIFICATE OF ANALYSIS


Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

Date Received : 30-Jun-10  
Date Tested : 07-Jul-10  
Date Completed : 21-Jul-10  
Purchase Order No. : 15565

Test Article : ClearKlens Tego 2000SS Block 2 VH25  
Sample Code :  
Batch Reference : W028159  
Laboratory Reference : LR-00972151 Article Quantity : 20 Bottles  
Test Required : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM03

Day Book Ref: 900620  
Volume Tested: 20x50mL  
Test carried out according to current Pharmacopoeias.



*Suz Wood*

Date *23 July 2010*

Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Laboratory Manager - Pharmaceutical Microbiology

*K. Grogan*

Sample Booking Number: CN\00152116  
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*