

Contact address:

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 23.07.10

Product Name	CLEARKLENS TEGO 2000SC VH25
Product Code	7513387
Batch Number	NO 28269
Production Date	07/2010
Expiry Date	07/2012

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	clear slightly yellow liquid
pH 100% (20°C)	JMD001	7.5 - 8.5	7.5
Specific Gravity (20°C)	JMD004	0.993 - 1.003	0.998

Signature / name *aga Fijvel*
On behalf of Manufacturing Location



FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P015645
Date Rec'd 15/07/10
Date: 19/07/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B210070535	DIVERSEY TEGO 2000SC 50ml 25-42kGy	104	46 X B/No.W028268 - CUST B/NO 324762 5 X B/No.W028267 - CUST B/NO 309113 53 C B/No.W028269 - CUST B/NO 327583 ACTUAL DOSE....33.3 - 33.5....kGy
B2FLE0010999 B210070535	SAMPLES	4	1 X B/No.W028268 1 X B/No.W028267 2 X B/No.W028269
Despatch Text: 3PLTS			
Total		108	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED