

Contact address:

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 16/8/10

Product Name	CLEARKLENS TEGO 2001SS VH24S
Product Code	7513403
Batch Number	W028304
Production Date	07/2010
Expiry Date	05/2011

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Liquid	clear liquid
pH 100% (20°C)	JDM001	8.5 - 9.2	8.5
Specific Gravity (20°C)	JDM004	0.998 - 1.001	0.998

Signature / name *C. Pascoe*
On behalf of Manufacturing Location



FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P015729
 Date Rec'd 5/08/10
 Date: 10/08/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010093 B210080234	TEGO 2001 SS 25-42kGy	132	B/No.W028304 - CUSTOMER B/No.7515138 ACTUAL DOSE....38.5 - 39.5....kGy
B2FLE0010999 B210080234	SAMPLES	1	CUSTOMERS BATCH/CODE No.s AS ORDER No.P015729 DATED 4/8/10
Despatch Text: 3PLTS			
Total		133	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED