

JohnsonDiversey

clean is just the beginning



Contact address:

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 6.09.10

Product Name	CLEARKLENS IPA VH1
Product Code	7513400
Batch Number	NO 28372
Production Date	2010/08
Expiry Date	2012/08

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Liquid	clear liquid
Specific Gravity (20°C)	JDM004	0.872 - 0.883	0.874

Signature / name *Olga Pijel*
On behalf of Manufacturing Location



FLE001

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited
Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P015781
Date Rec'd 19/08/10
Date: 25/08/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B210080576	JD IPA BOTTLES 25-45kGy	432	B/No.W028372 ACTUAL DOSE...37.6 - 38.8...kGy CUSTOMERS BATCH/CODE No.s AS ORDER No.P015781 DATED 19/8/10
	Despatch Text: 8PLTS		
	Total	432	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS
 Cust. Ref: P015821
 Date Rec'd 26/08/10
 Date: 1/09/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B210080793	JD IPA BOTTLES 25-45kGy	281	W028372 ACTUAL DOSE...37.1 - 39.1....kGy
B2FLE0010999 B210080793	SAMPLES	5	CUSTOMERS BATCH/CODE NO'S AS ORDER NO P015821 DATED 25.08.10
Despatch Text: 27TUBS			
Total		286	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 07-Sep-10
Date Tested : 13-Sep-10
Date Completed : 13-Sep-10
Purchase Order No. : 15842

Test Article : Clearklens IPA Block 1
Sample Code : Works Order No. 28372
Batch Reference : 1985
Laboratory Reference : LR-00982367 **Article Quantity** : 1xBottle
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. Ph Eur and USP.

Test	Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	* MM110	<0.2000 EU/mL
Dilution	* MM110/01	1/40
Spiked Recoveries	* MM110/02	Valid 81%
Endotoxin Spike Level	* MM110/09	0.5 EU/mL
Product Standard Data Sheet	* MQ002/01	FM02

Test carried out according to current Pharmacopoeias.

John Jones 20/09/10

Suz Wood

Date *17 Sep 2010*



Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN00154143

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS


Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 07-Sep-10
Date Tested : 09-Sep-10
Date Completed : 23-Sep-10
Purchase Order No. : 15842

Test Article : Clearklens IPA Block 1
Sample Code : Works Order No. 28372
Batch Reference : 1985
Laboratory Reference : LR-00982366 **Article Quantity** : 20xBottles
Test Required : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM03

Day Book Ref: 000011
Volume Tested: 20x50mL
Test carried out according to current Pharmacopoeias.



Mr K.A. Barber, CBiol, MSB
Business Manager

Date 22 Sep 2010

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN00154143
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited