

Remove most of water
Case lost off with Filter

JohnsonDiversey
Clean is just the beginning



Diversey

Contact

JohnsonDiversey Europe BV
Maarssenbroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 7/2/11

Product Name	CLEARLENs BI-SPORE ACTIVATOR SOLUTION VH26
Product Code	7514852
Batch Number	WO 29329
Production Date	02/2011
Expiry Date	02/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Colourless Liquid	clear colourless liquid
pH (100% 20°C)	JDM001	11.5 - 13.5	11.5
Specific Gravity (20°C)	JDM004	1.015 - 1.035	1.017

"SIGNATURE"

C. Pascoe

Name

Caroline Pascoe

Function Name

Pharmaceutical Supervisor

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

ACTIVATOR 10ST OFF WITH FILTER
Base 10ST OFF WITH FILTER

JohnsonDiversey
Clean is just the beginning



Diversey

Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 8 02. 2011

Product Name	CLEARLENs BI-SPORE BASE SOLUTION VH26
Product Code	7514852
Batch Number	W029329
Production Date	02/2011
Expiry Date	02/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Blue Liquid	clear blue liquid
pH (100% 20°C)	JDM001	1.5 - 3.5	1.9
Specific Gravity (20°C)	JDM004	1.020 - 1.060	1.025

"SIGNATURE"

Maria

Name *Gunnishe*
Function Name *Pharmo Supervisor*

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

0476207 00

FLE001

LA1 4X5



Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

All in accordance with current
technical agreement

Authorising Signature:

For and on behalf of ISOTRON LIMITED



Wickham Laboratories
contract analytical services

Wickham Laboratories Limited,
Winchester Road, Wickham, Fareham,
Hampshire PO17 5EU England
Telephone: +44(0)1329 832511
Fax: +44(0)1329 834262

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 09 Feb 11
Date Tested: 15 Feb 11
Date Completed: 01 Mar 11
Purchase Order: 16532

Test Article: Clarklens Bi Spore
Sample Code: Base Fluid
Batch Ref: WO29329
Qty Received: 20 x 100 mL
Laboratory Reference: 0001083/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Product Standard Data Sheet	MM107/03	FM05
Volume Tested	MM107/05	20 x 50 mL

External Comments

Test carried out according to current Pharmacopocias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 02 Mar 11 15:36:26

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0001083
Lancaster

Certificate of Analysis - OSMM

Print Number

P0002842

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



Wickham Laboratories
contract analytical services

Wickham Laboratories Limited,
Winchester Road, Wickham, Fareham,
Hampshire PO17 5EU England
Telephone: +44(0)1329 832511
Fax: +44(0)1329 834262

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 09 Feb 11
Date Tested: 15 Feb 11
Date Completed: 01 Mar 11
Purchase Order: 16532

Test Article: Clarklens Bi Spore
Sample Code: Activator Fluid
Batch Ref: WO29329
Qty Received: 20 x 100 mL
Laboratory Reference: 0001083/2
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Product Standard Data Sheet	MM107/03	FM04
Volume Tested	MM107/05	20 x 50 mL

External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 02 Mar 11 15:36:26

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0001083
Lancaster

Certificate of Analysis - OSMM

Print Number

P0002842

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited