

Remove last off with filter

JohnsonDiversey
Clean is just the beginning



Diversey

Contact

JohnsonDiversey Europe BV
Maarssenbroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 15/2/11

Product Name	CLEARLENs BI-SPORE ACTIVATOR SOLUTION VH26
Product Code	7514852
Batch Number	4029363
Production Date	02/2011
Expiry Date	02/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Colourless Liquid	clear Colourless liquid
pH (100% 20°C)	JDM001	11.5 - 13.5	11.5
Specific Gravity (20°C)	JDM004	1.015 - 1.035	1.017

"SIGNATURE" C. Pascoe

Name Caroline Pascoe
Function Name Phamroom Supervisor

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

Remove 100% of water
Base last off with filter

JohnsonDiversey
Clean is just the beginning



Diversey

Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 15/2/11

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26
Product Code	7514852
Batch Number	4029363
Production Date	02/2011
Expiry Date	02/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Blue Liquid	clear blue liquid
pH (100% 20°C)	JDM001	1.5 - 3.5	1.9
Specific Gravity (20°C)	JDM004	1.020 - 1.060	1.025

"SIGNATURE"

C. P. P. P.

Name

Caroline P. P. P.

Function Name

Pharma Research Supervisor

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

Activate the unit with filter
Base lost off with filter

JohnsonDiversey
Clean is just the beginning



Diversey

Contact

JohnsonDiversey Europe BV
Maarssenbroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 15/2/11

Product Name	CLEARKLENS BI-SPORE COMBINED SOLUTION VH26
Product Code	7514852
Batch Number	4029363
Production Date	02/2011
Expiry Date	02/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Yellow Liquid	clear yellow liquid
pH (100% 20°C)	JDM001	2.5 - 5.5	2.9
Specific Gravity (20°C)	JDM004	1.000 - 1.010	1.001
Potential Chlorine Dioxide		230 - 280ppm	250ppm

"SIGNATURE"

C. Alcoe

Name

Caroline Alcoe

Function Name

Pharmaceutical Supervisor

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P016500
Date Rec'd 3/02/11
Date: 9/02/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010113 B211020171	JD-BOTTLES/CAPS/BAGS 25-40kGy	8	B/No.W029362 / W029363. CUSTOMERS BATCH/CODE No.s AS ORDER No.P016500 DATED 2/2/11
B2FLE001MIN	MINIMUM CHARGE UPLIFT		
Despatch Text: 2PLTS			
Total		8	Last Page 1 of 1

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 17 Feb 11
Date Tested: 22 Feb 11
Date Completed: 08 Mar 11
Purchase Order: P016587

Test Article: ClearKlens Bi Spore Base
Sample Code: 101302
Batch Ref: WO29363
Qty Received: 20 x 100 mL
Laboratory Reference: 0001133/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Product Standard Data Sheet	MM107/03	FM05
Volume Tested	MM107/05	20 x 50 mL

External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 09 Mar 11 16:33:47

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0001133
Lancaster

Certificate of Analysis - OSMM

Print Number

P0002908

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



Wickham Laboratories
contract analytical services

Wickham Laboratories Limited,
Winchester Road, Wickham, Fareham,
Hampshire PO17 5EU England
Telephone: +44(0)1329 832511
Fax: +44(0)1329 834262

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 17 Feb 11
Date Tested: 22 Feb 11
Date Completed: 08 Mar 11
Purchase Order: P016587

Test Article: ClearKlens Bi Spore Activator
Sample Code: 101303
Batch Ref: WO29363
Qty Received: 20 x 100 mL
Laboratory Reference: 0001133/2
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Product Standard Data Sheet	MM107/03	FM04
Volume Tested	MM107/05	20 x 50 mL

External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 09 Mar 11 16:33:47

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0001133
Lancaster

Certificate of Analysis - OSMM

Print Number

P0002908

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited